

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111218\
 Data File : BE098189.D
 Acq On : 12 Nov 2018 14:13
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE111218

Quant Time: Nov 12 14:43:48 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 12 13:32:15 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	103	0.00
2	1,4-Dioxane	0.400	0.436	-9.0	117	0.00
3	n-Nitrosodimethylamine	0.400	0.358	10.5	99	0.00
4 S	2-Fluorophenol	0.400	0.397	0.8	104	0.00
5 S	Phenol-d6	0.400	0.371	7.3	98	0.00
6	bis(2-Chloroethyl)ether	0.400	0.366	8.5	100	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	102	0.00
8 S	Nitrobenzene-d5	0.400	0.401	-0.3	108	0.01
9	Naphthalene	0.400	0.388	3.0	102	0.00
10	Hexachlorobutadiene	0.400	0.372	7.0	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.387	3.3	102	0.00
12	2-Methylnaphthalene	0.400	0.393	1.8	104	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.400	0.339	15.3	102	0.01
15 S	2-Fluorobiphenyl	0.400	0.381	4.8	105	0.00
16	Acenaphthylene	0.400	0.394	1.5	106	0.00
17	Acenaphthene	0.400	0.392	2.0	105	0.00
18	Fluorene	0.400	0.389	2.8	104	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	102	0.00
20	4-Bromophenyl-phenylether	0.400	0.396	1.0	108	0.01
21	Hexachlorobenzene	0.400	0.386	3.5	102	0.01
22	Pentachlorophenol	0.400	0.366	8.5	101	0.00
23	Phenanthrene	0.400	0.388	3.0	103	0.01
24	Anthracene	0.400	0.382	4.5	102	0.01
25 SURR	Fluoranthene-d10	0.400	0.375	6.3	100	0.00
26	Fluoranthene	0.400	0.383	4.3	102	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	100	0.00
28	Pvrene	0.400	0.369	7.8	101	0.00
29 S	Terphenyl-d14	0.400	0.357	10.8	99	0.00
30	Benzo(a)anthracene	0.400	0.370	7.5	102	0.00
31	Chrysene	0.400	0.371	7.3	101	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.345	13.8	98	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.379	5.3	100	0.00
34 I	Perylene-d12	0.400	0.400	0.0	100	0.00
35	Benzo(b)fluoranthene	0.400	0.368	8.0	99	0.00
36	Benzo(k)fluoranthene	0.400	0.386	3.5	103	0.00
37 C	Benzo(a)pyrene	0.400	0.378	5.5	101	0.00
38	Dibenzo(a,h)anthracene	0.400	0.370	7.5	98	0.01
39	Benzo(a,h,i)perylene	0.400	0.379	5.3	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111218\
Data File : BE098189.D
Acq On : 12 Nov 2018 14:13
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE111218

Quant Time: Nov 12 14:43:48 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 12 13:32:15 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				