

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111218\
 Data File : BE098206.D
 Acq On : 13 Nov 2018 1:37
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 13 02:08:17 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 12 13:32:15 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.650	0.569	12.5	100	0.00
3	n-Nitrosodimethylamine	0.785	0.724	7.8	100	0.00
4 S	2-Fluorophenol	1.129	1.124	0.4	103	0.00
5 S	Phenol-d6	1.814	1.797	0.9	103	0.00
6	bis(2-Chloroethyl)ether	1.464	1.356	7.4	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.399	0.391	2.0	103	0.00
9	Naphthalene	4.068	3.923	3.6	99	0.00
10	Hexachlorobutadiene	0.253	0.233	7.9	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.702	0.684	2.6	100	0.00
12	2-Methylnaphthalene	0.706	0.698	1.1	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.174	0.158	9.2	112	0.00
15 S	2-Fluorobiphenyl	1.675	1.528	8.8	103	0.00
16	Acenaphthylene	1.834	1.769	3.5	106	0.00
17	Acenaphthene	1.120	1.083	3.3	107	0.00
18	Fluorene	1.434	1.385	3.4	106	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	4-Bromophenyl-phenylether	0.244	0.239	2.0	109	0.00
21	Hexachlorobenzene	0.254	0.246	3.1	105	0.01
22	Pentachlorophenol	0.044	0.052	-18.2	134	0.00
23	Phenanthrene	0.996	0.970	2.6	106	0.00
24	Anthracene	0.928	0.909	2.0	107	0.00
25 SURR	Fluoranthene-d10	4.995	4.846	3.0	106	0.00
26	Fluoranthene	1.234	1.204	2.4	107	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
28	Pvrene	1.101	1.052	4.5	109	0.00
29 S	Terphenyl-d14	0.908	0.840	7.5	106	0.00
30	Benzo(a)anthracene	1.137	1.085	4.6	109	0.00
31	Chrysene	1.154	1.097	4.9	107	0.00
32	Bis(2-ethylhexyl)phthalate	2.360	2.430	-3.0	122	0.00
33	Indeno(1,2,3-cd)pyrene	1.058	1.075	-1.6	111	0.00
34 I	Perylene-d12	1.000	1.000	0.0	104	0.00
35	Benzo(b)fluoranthene	1.135	1.076	5.2	108	0.00
36	Benzo(k)fluoranthene	1.197	1.134	5.3	106	0.00
37 C	Benzo(a)pyrene	1.088	1.039	4.5	107	0.00
38	Dibenzo(a,h)anthracene	0.927	0.944	-1.8	114	0.00
39	Benzo(a,h,i)perylene	0.968	0.929	4.0	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111218\
Data File : BE098206.D
Acq On : 13 Nov 2018 1:37
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 13 02:08:17 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 12 13:32:15 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			