

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111418\
 Data File : BE098224.D
 Acq On : 14 Nov 2018 14:21
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 14 15:16:57 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 12 13:32:15 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.650	0.578	11.1	102	-0.01
3	n-Nitrosodimethylamine	0.785	0.748	4.7	103	0.00
4 S	2-Fluorophenol	1.129	1.123	0.5	103	0.00
5 S	Phenol-d6	1.814	1.777	2.0	102	0.00
6	bis(2-Chloroethyl)ether	1.464	1.423	2.8	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.399	0.409	-2.5	108	0.00
9	Naphthalene	4.068	4.149	-2.0	105	0.00
10	Hexachlorobutadiene	0.253	0.246	2.8	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.702	0.712	-1.4	105	0.00
12	2-Methylnaphthalene	0.706	0.714	-1.1	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.174	0.157	9.8	104	0.00
15 S	2-Fluorobiphenyl	1.675	1.632	2.6	103	0.00
16	Acenaphthylene	1.834	1.881	-2.6	106	0.00
17	Acenaphthene	1.120	1.146	-2.3	106	0.00
18	Fluorene	1.434	1.464	-2.1	105	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4-Bromophenyl-phenylether	0.244	0.247	-1.2	103	0.00
21	Hexachlorobenzene	0.254	0.254	0.0	99	0.00
22	Pentachlorophenol	0.044	0.046	-4.5	107	0.00
23	Phenanthrene	0.996	1.027	-3.1	102	0.00
24	Anthracene	0.928	0.939	-1.2	101	0.00
25 SURR	Fluoranthene-d10	4.995	4.947	1.0	99	0.00
26	Fluoranthene	1.234	1.244	-0.8	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
28	Pvrene	1.101	1.162	-5.5	102	0.00
29 S	Terphenyl-d14	0.908	0.929	-2.3	100	0.00
30	Benzo(a)anthracene	1.137	1.186	-4.3	101	0.00
31	Chrysene	1.154	1.241	-7.5	103	0.00
32	Bis(2-ethylhexyl)phthalate	2.360	2.484	-5.3	106	0.00
33	Indeno(1,2,3-cd)pyrene	1.058	1.082	-2.3	95	0.00
34 I	Perylene-d12	1.000	1.000	0.0	91	0.00
35	Benzo(b)fluoranthene	1.135	1.148	-1.1	100	0.00
36	Benzo(k)fluoranthene	1.197	1.274	-6.4	104	0.00
37 C	Benzo(a)pyrene	1.088	1.105	-1.6	99	0.00
38	Dibenzo(a,h)anthracene	0.927	0.902	2.7	95	0.00
39	Benzo(a,h,i)perylene	0.968	0.976	-0.8	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111418\
Data File : BE098224.D
Acq On : 14 Nov 2018 14:21
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 14 15:16:57 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 12 13:32:15 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			