

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE111419\
 Data File : BE100727.D
 Acq On : 14 Nov 2019 18:54
 Operator : JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE111419

Quant Time: Nov 14 19:44:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 14 18:50:05 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.697	0.780	-11.9	127	0.00
3	n-Nitrosodimethylamine	0.692	0.739	-6.8	122	0.00
4 S	2-Fluorophenol	1.168	1.141	2.3	106	0.00
5 S	Phenol-d6	1.652	1.600	3.1	107	0.00
6	bis(2-Chloroethyl)ether	1.444	1.391	3.7	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.340	0.330	2.9	106	0.00
9	Naphthalene	4.146	4.223	-1.9	104	0.00
10	Hexachlorobutadiene	0.192	0.200	-4.2	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.638	0.639	-0.2	101	0.00
12	2-Methylnaphthalene	0.707	0.718	-1.6	106	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	-0.01
14 S	2,4,6-Tribromophenol	0.042	0.031	26.2#	106	-0.01
15 S	2-Fluorobiphenyl	1.467	1.532	-4.4	103	0.00
16	Acenaphthylene	1.617	1.538	4.9	100	0.00
17	Acenaphthene	1.063	1.078	-1.4	104	0.00
18	Fluorene	1.453	1.462	-0.6	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4-Bromophenyl-phenylether	0.211	0.205	2.8	99	0.00
21	Hexachlorobenzene	0.205	0.206	-0.5	104	0.00
22	Pentachlorophenol	0.000	0.000	0.0	0#	0.03
23	Phenanthrene	1.090	1.119	-2.7	104	0.00
24	Anthracene	0.974	0.942	3.3	103	0.01
25 SURR	Fluoranthene-d10	5.287	5.199	1.7	101	0.00
26	Fluoranthene	1.456	1.426	2.1	102	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
28	Pvrene	1.112	1.098	1.3	101	0.00
29 S	Terphenyl-d14	0.954	0.949	0.5	102	0.00
30	Benzo(a)anthracene	1.284	1.266	1.4	101	0.00
31	Chrysene	1.303	1.302	0.1	101	0.00
32	Bis(2-ethylhexyl)phthalate	2.278	1.836	19.4	92	0.00
33	Indeno(1,2,3-cd)pyrene	1.527	1.511	1.0	99	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.167	1.174	-0.6	100	0.00
36	Benzo(k)fluoranthene	1.223	1.245	-1.8	102	0.00
37 C	Benzo(a)pyrene	1.088	1.096	-0.7	100	0.00
38	Dibenzo(a,h)anthracene	1.134	1.126	0.7	99	0.00
39	Benzo(a,h,i)perylene	1.113	1.132	-1.7	101	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE111419\
Data File : BE100727.D
Acq On : 14 Nov 2019 18:54
Operator : JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE111419

Quant Time: Nov 14 19:44:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 14 18:50:05 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			