

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE111519\
 Data File : BE100741.D
 Acq On : 15 Nov 2019 17:31
 Operator : JU
 Sample : K5728-05
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PST-024-SW135-110619-1

Quant Time: Nov 15 22:33:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 15 15:55:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	3437	0.40	ng	0.00
7) Naphthalene-d8	10.55	136	14910	0.40	ng	0.00
13) Acenaphthene-d10	14.40	164	9718	0.40	ng	-0.01
19) Phenanthrene-d10	17.14	188	22214	0.40	ng	0.00
27) Chrysene-d12	21.29	240	24050	0.40	ng	0.00
34) Perylene-d12	23.73	264	25613	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.37	112	2584	0.26	ng	0.00
5) Phenol-d6	6.94	99	4258	0.30	ng	0.00
8) Nitrobenzene-d5	8.91	82	2847	0.22	ng	0.00
11) 2-Methylnaphthalene-d10	12.15	152	5400	0.23	ng	0.00
14) 2,4,6-Tribromophenol	15.89	330	311	0.43	ng	0.00
15) 2-Fluorobiphenyl	13.03	172	7136	0.20	ng	0.00
25) Fluoranthene-d10	19.16	212	56891	0.19	ng	0.00
29) Terphenyl-d14	19.75	244	10700	0.19	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Naphthalene	10.60	128	7793	0.050	ng	# 97
12) 2-Methylnaphthalene	12.22	142	1873	0.071	ng	97
16) Acenaphthylene	14.13	152	16847	0.429	ng	98
17) Acenaphthene	14.47	154	1035	0.040	ng	# 88
18) Fluorene	15.43	166	4589	0.130	ng	# 82
23) Phenanthrene	17.18	178	108143	1.786	ng	99
24) Anthracene	17.26	178	8949	0.165	ng	96
26) Fluoranthene	19.18	202	150264	1.858	ng	98
28) Pyrene	19.55	202	211961	3.170	ng	97
30) Benzo(a)anthracene	21.33	228	105762	1.370	ng	94
31) Chrysene	21.33	228	105126	1.342	ng	98
32) Bis(2-ethylhexyl)phthalate	21.22	149	27883	0.204	ng	98
33) Indeno(1,2,3-cd)pyrene	26.29	276	51797	0.564	ng	97
35) Benzo(b)fluoranthene	22.99	252	99704	1.334	ng	100
36) Benzo(k)fluoranthene	22.99	252	99704	1.273	ng	99
37) Benzo(a)pyrene	23.62	252	75917	1.090	ng	99
38) Dibenzo(a,h)anthracene	26.30	278	13581	0.187	ng	# 89
39) Benzo(a,h,i)perylene	27.08	276	59023	0.828	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE111519\
Data File : BE100741.D
Acq On : 15 Nov 2019 17:31
Operator : JU
Sample : K5728-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PST-024-SW135-110619-1

Quant Time: Nov 15 22:33:07 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 15 15:55:40 2019
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 0.400 ng

RT: 7.77 min Scan# 426

Delta R.T. -0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31

Instrument :

BNA_E

ClientSampleId :

PST-024-SW135-110619-1


Tot Ion:152 Resp: 3437

Ion Ratio Lower Upper

152 100

150 158.8 123.4 185.0




115 64.9 52.1 78.1



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#4

2-Fluorophenol

Concen: 0.257 ng

RT: 5.37 min Scan# 217

Delta R.T. -0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31


Tgt Ion:112 Resp: 2584

Ion Ratio Lower Upper

112 100

64 71.8 58.3 87.5



63 151.7 115.4 173.2




Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BE1

Ion 63.00 (62.70 to 63.70): BE1





#5

Phenol-d6

Concen: 0.300 ng

RT: 6.94 min Scan# 354

Delta R.T. -0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31

Instrument :

BNA_E

ClientSampleId :

PST-024-SW135-110619-1


Tot Ion: 99 Resp: 4258

Ion Ratio Lower Upper




99 100

42 29.8 17.8 26.6#

71 33.1 27.3 40.9



Abundance Ion 99.00 (98.70 to 99.70): BE1
Ion 42.00 (41.70 to 42.70): BE1
Ion 71.00 (70.70 to 71.70): BE1



#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.55 min Scan# 645

Delta R.T. 0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31

Tgt Ion: 136 Resp: 14910


Ion Ratio Lower Upper

136 100



137 11.6 9.1 13.7


54 44.8 39.0 58.6

68 8.8 7.8 11.8



Abundance Ion 136.00 (135.70 to 136.70): BE1
Ion 137.00 (136.70 to 137.70): BE1
Ion 54.00 (53.70 to 54.70): BE1
Ion 68.00 (67.70 to 68.70): BE1









#8
 Nitrobenzene-d5
 Concen: 0.225 ng
 RT: 8.91 min Scan# 519
 Delta R.T. 0.00 min
 Lab File: BE100741.D
 Acq: 15 Nov 2019 17:31

Instrument :
 BNA_E
 ClientSampleId :
 PST-024-SW135-110619-1

Tot Ion: 82 Resp: 2847
 Ion Ratio Lower Upper
 82 100
 128 129.7 99.6 149.4
 54 239.7 183.8 275.8




Abundance Ion 82.00 (81.70 to 82.70): BE1
 Ion 128.00 (127.70 to 128.70): BE1
 Ion 54.00 (53.70 to 54.70): BE1






#9
 Naphthalene
 Concen: 0.050 ng
 RT: 10.60 min Scan# 649
 Delta R.T. 0.00 min
 Lab File: BE100741.D
 Acq: 15 Nov 2019 17:31

Tgt Ion: 128 Resp: 7793
 Ion Ratio Lower Upper
 128 100
 129 4.0 2.4 3.6#
 127 4.2 2.7 4.1#



Abundance Ion 128.00 (127.70 to 128.70): E
 Ion 129.00 (128.70 to 129.70): E
 Ion 127.00 (126.70 to 127.70): E









#11
2-Methylnaphthalene-d10
Concen: 0.227 ng
RT: 12.15 min Scan# 759
Delta R.T. 0.00 min
Lab File: BE100741.D
Acq: 15 Nov 2019 17:31




Instrument :
BNA_E
ClientSampleId :
PST-024-SW135-110619-1


Tot Ion:152 Resp: 5400
Ion Ratio Lower Upper
152 100
151 19.3 14.2 21.4



#12
2-Methylnaphthalene
Concen: 0.071 ng
RT: 12.22 min Scan# 764
Delta R.T. 0.00 min
Lab File: BE100741.D
Acq: 15 Nov 2019 17:31

Tgt Ion:142 Resp: 1873
Ion Ratio Lower Upper
142 100
141 88.9 73.3 109.9
115 39.7 29.9 44.9





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.40 min Scan# 915

Delta R.T. -0.01 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31

Instrument :

BNA_E

ClientSampleId :

PST-024-SW135-110619-1


Tot Ion:164 Resp: 9718

Ion Ratio Lower Upper

164 100

162 100.2 77.8 116.8




160 47.9 35.3 52.9



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#14

2,4,6-Tribromophenol

Concen: 0.426 ng

RT: 15.89 min Scan# 1020

Delta R.T. 0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31


Tgt Ion:330 Resp: 311

Ion Ratio Lower Upper

330 100

332 96.5 85.0 127.6



141 108.7 28.2 42.2#




Abundance Ion 330.00 (329.70 to 330.70): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#15
2-Fluorobiphenyl
Concen: 0.200 ng
RT: 13.03 min Scan# 820
Delta R.T. 0.00 min
Lab File: BE100741.D
Acq: 15 Nov 2019 17:31

Instrument :
BNA_E
ClientSampleId :
PST-024-SW135-110619-1

Tot Ion:172 Resp: 7136
Ion Ratio Lower Upper
172 100
171 34.7 28.1 42.1
170 23.9 19.8 29.6




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#16
Acenaphthylene
Concen: 0.429 ng
RT: 14.13 min Scan# 896
Delta R.T. 0.00 min
Lab File: BE100741.D
Acq: 15 Nov 2019 17:31

Tgt Ion:152 Resp: 16847
Ion Ratio Lower Upper
152 100
151 19.8 14.9 22.3
153 14.1 10.7 16.1



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E





#17

Acenaphthene

Concen: 0.040 ng

RT: 14.47 min Scan# 920

Delta R.T. 0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31

Instrument :

BNA_E

ClientSampleId :

PST-024-SW135-110619-1


Tot Ion:154 Resp: 1035

Ion Ratio Lower Upper

154 100

153 125.4 91.9 137.9




152 68.7 44.8 67.2#



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18

Fluorene

Concen: 0.130 ng

RT: 15.43 min Scan# 986

Delta R.T. 0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31


Tgt Ion:166 Resp: 4589

Ion Ratio Lower Upper

166 100

165 115.0 78.2 117.4



167 20.9 10.6 16.0#




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.14 min Scan# 1113

Delta R.T. 0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31

Instrument :

BNA_E

ClientSampleId :

PST-024-SW135-110619-1


Tot Ion:188 Resp: 22214

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0




80 5.7 5.4 8.0



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE1

Ion 80.00 (79.70 to 80.70): BE1



#23

Phenanthrene

Concen: 1.786 ng

RT: 17.18 min Scan# 1116

Delta R.T. 0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31


Tgt Ion:178 Resp: 108143

Ion Ratio Lower Upper

178 100

176 19.4 15.0 22.4



179 15.6 12.6 18.8




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#24

Anthracene

Concen: 0.165 ng

RT: 17.26 min Scan# 1122

Delta R.T. 0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31

Instrument :

BNA_E

ClientSampleId :

PST-024-SW135-110619-1




Tot Ion:178 Resp: 8949

Ion Ratio Lower Upper

178 100

176 17.1 14.6 21.8

179 13.5 12.6 18.8



#25

Fluoranthene-d10

Concen: 0.194 ng

RT: 19.16 min Scan# 1373

Delta R.T. -0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31




Tgt Ion:212 Resp: 56891


Ion Ratio Lower Upper

212 100

106 2.8 2.2 3.4

104 1.6 1.2 1.8






#26
 Fluoranthene
 Concen: 1.858 ng
 RT: 19.18 min Scan# 1378
 Delta R.T. -0.00 min
 Lab File: BE100741.D
 Acq: 15 Nov 2019 17:31

Instrument :
 BNA_E
 ClientSampleId :
 PST-024-SW135-110619-1

Tot Ion	Ratio	Lower	Upper
202	100		
101	10.3	8.6	12.8
203	18.0	13.6	20.4






Abundance

Ion 202.00 (201.70 to 202.70): E


Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#27
 Chrysene-d12
 Concen: 0.400 ng
 RT: 21.29 min Scan# 1624
 Delta R.T. 0.00 min
 Lab File: BE100741.D
 Acq: 15 Nov 2019 17:31

Tgt Ion	Ratio	Lower	Upper
240	100		
120	7.9	7.4	11.0
236	28.1	21.8	32.6






Abundance

Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#28

Pvrene

Concen: 3.170 ng

RT: 19.55 min Scan# 1441

Delta R.T. -0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31

Instrument :

BNA_E

ClientSampleId :

PST-024-SW135-110619-1


Tot Ion:202 Resp: 211961

Ion Ratio Lower Upper

202 100

200 21.1 17.0 25.4

203 19.9 14.0 21.0






Abundance

Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#29

Terphenyl-d14

Concen: 0.187 ng

RT: 19.75 min Scan# 1477

Delta R.T. -0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31


Tgt Ion:244 Resp: 10700

Ion Ratio Lower Upper

244 100

212 37.0 24.3 36.5#

122 10.7 7.6 11.4






Abundance

Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#30

Benzo(a)anthracene

Concen: 1.370 ng

RT: 21.33 min Scan# 1627

Delta R.T. 0.05 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31

Instrument :

BNA_E

ClientSampleId :

PST-024-SW135-110619-1


Tot Ion:228 Resp: 105762

Ion Ratio Lower Upper

228 100

226 30.5 21.1 31.7

229 21.8 15.9 23.9






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#31

Chrysene

Concen: 1.342 ng

RT: 21.33 min Scan# 1627

Delta R.T. 0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31


Tgt Ion:228 Resp: 105126

Ion Ratio Lower Upper

228 100

226 30.5 24.5 36.7

229 21.8 15.5 23.3






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#32

Bis(2-ethylhexyl)phthalate

Concen: 0.204 ng

RT: 21.22 min Scan# 1618

Delta R.T. 0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31

Instrument :

BNA_E

ClientSampleId :

PST-024-SW135-110619-1


Tot Ion:149 Resp: 27883

Ion Ratio Lower Upper

149 100

167 5.8 5.2 7.8




279 0.9 0.9 1.3



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#33

Indeno(1,2,3-cd)pyrene

Concen: 0.564 ng

RT: 26.29 min Scan# 2875

Delta R.T. -0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31


Tgt Ion:276 Resp: 51797

Ion Ratio Lower Upper

276 100

138 15.7 14.8 22.2



277 24.7 19.6 29.4




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#34
Pervlene-d12
Concen: 0.400 ng
RT: 23.73 min Scan# 2161
Delta R.T. 0.00 min
Lab File: BE100741.D
Acq: 15 Nov 2019 17:31




Instrument :
BNA_E
ClientSampleId :
PST-024-SW135-110619-1


Tot Ion:264 Resp: 25613
Ion Ratio Lower Upper
264 100
260 24.3 19.1 28.7
265 28.2 19.7 29.5



#35
Benzo(b)fluoranthene
Concen: 1.334 ng
RT: 22.99 min Scan# 1953
Delta R.T. 0.00 min
Lab File: BE100741.D
Acq: 15 Nov 2019 17:31

Tgt Ion:252 Resp: 99704
Ion Ratio Lower Upper
252 100
253 23.2 18.7 28.1
125 8.8 7.0 10.4









#36
Benzo(k)fluoranthene
Concen: 1.273 ng
RT: 22.99 min Scan# 1953
Delta R.T. -0.05 min
Lab File: BE100741.D
Acq: 15 Nov 2019 17:31




Instrument :
BNA_E
ClientSampleId :
PST-024-SW135-110619-1


Tot Ion:252 Resp: 99704
Ion Ratio Lower Upper
252 100
253 23.2 18.6 27.8
125 8.8 7.6 11.4



#37
Benzo(a)pyrene
Concen: 1.090 ng
RT: 23.62 min Scan# 2129
Delta R.T. -0.00 min
Lab File: BE100741.D
Acq: 15 Nov 2019 17:31

Tgt Ion:252 Resp: 75917
Ion Ratio Lower Upper
252 100
253 23.0 18.6 28.0
125 9.1 7.9 11.9





#38

Dibenzo(a,h)anthracene

Concen: 0.187 ng

RT: 26.30 min Scan# 2878

Delta R.T. -0.01 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31

Instrument :

BNA_E

ClientSampleId :

PST-024-SW135-110619-1


Tot Ion:278 Resp: 13581

Ion Ratio Lower Upper




278 100

139 16.8 10.3 15.5#

279 31.9 20.6 31.0#



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#39

Benzo(g,h,i)perylene

Concen: 0.828 ng

RT: 27.08 min Scan# 3095

Delta R.T. -0.00 min

Lab File: BE100741.D

Acq: 15 Nov 2019 17:31


Tgt Ion:276 Resp: 59023

Ion Ratio Lower Upper

276 100

277 24.5 19.8 29.6

138 16.8 13.6 20.4



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

