

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE111519\
 Data File : BE100729.D
 Acq On : 15 Nov 2019 10:03
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 15 22:30:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 15 15:55:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.697	0.739	-6.0	142	0.00
3	n-Nitrosodimethylamine	0.692	0.679	1.9	132	0.00
4 S	2-Fluorophenol	1.168	1.156	1.0	127	0.00
5 S	Phenol-d6	1.652	1.576	4.6	124	0.00
6	bis(2-Chloroethyl)ether	1.444	1.427	1.2	125	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	126	0.00
8 S	Nitrobenzene-d5	0.340	0.313	7.9	121	0.00
9	Naphthalene	4.146	4.238	-2.2	126	0.00
10	Hexachlorobutadiene	0.192	0.191	0.5	126	0.00
11 SURR	2-Methylnaphthalene-d10	0.638	0.642	-0.6	123	0.00
12	2-Methylnaphthalene	0.707	0.709	-0.3	127	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	-0.01
14 S	2,4,6-Tribromophenol	0.042	0.030	28.6#	121	-0.01
15 S	2-Fluorobiphenyl	1.467	1.523	-3.8	123	0.00
16	Acenaphthylene	1.617	1.495	7.5	116	0.00
17	Acenaphthene	1.063	1.062	0.1	123	0.00
18	Fluorene	1.453	1.435	1.2	121	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
20	4-Bromophenyl-phenylether	0.211	0.203	3.8	114	0.00
21	Hexachlorobenzene	0.205	0.196	4.4	114	0.00
22	Pentachlorophenol	0.000	0.000	0.0	0#	-0.01
23	Phenanthrene	1.090	1.084	0.6	117	0.00
24	Anthracene	0.974	0.904	7.2	114	0.00
25 SURR	Fluoranthene-d10	5.287	5.069	4.1	114	0.00
26	Fluoranthene	1.456	1.384	4.9	115	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	116	0.00
28	Pvrene	1.112	1.106	0.5	115	0.00
29 S	Terphenyl-d14	0.954	0.944	1.0	114	0.00
30	Benzo(a)anthracene	1.284	1.211	5.7	109	0.00
31	Chrysene	1.303	1.290	1.0	113	0.00
32	Bis(2-ethylhexyl)phthalate	2.278	1.598	29.9#	90	0.00
33	Indeno(1,2,3-cd)pyrene	1.527	1.498	1.9	111	0.00
34 I	Perylene-d12	1.000	1.000	0.0	114	0.00
35	Benzo(b)fluoranthene	1.167	1.150	1.5	111	0.00
36	Benzo(k)fluoranthene	1.223	1.173	4.1	109	0.00
37 C	Benzo(a)pyrene	1.088	1.057	2.8	110	0.00
38	Dibenzo(a,h)anthracene	1.134	1.114	1.8	111	0.00
39	Benzo(a,h,i)perylene	1.113	1.089	2.2	110	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE111519\
Data File : BE100729.D
Acq On : 15 Nov 2019 10:03
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 15 22:30:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 15 15:55:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			