

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE111519\
 Data File : BE100745.D
 Acq On : 15 Nov 2019 20:00
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 15 23:53:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 15 15:55:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	223#	0.00
2	1,4-Dioxane	0.697	0.600	13.9	199#	0.01
3	n-Nitrosodimethylamine	0.692	0.693	-0.1	234#	0.01
4 S	2-Fluorophenol	1.168	1.271	-8.8	241#	0.00
5 S	Phenol-d6	1.652	1.937	-17.3	264#	0.00
6	bis(2-Chloroethyl)ether	1.444	1.458	-1.0	221#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	233#	0.00
8 S	Nitrobenzene-d5	0.340	0.357	-5.0	257#	0.00
9	Naphthalene	4.146	4.085	1.5	226#	0.00
10	Hexachlorobutadiene	0.192	0.184	4.2	225#	0.00
11 SURR	2-Methylnaphthalene-d10	0.638	0.622	2.5	222#	0.00
12	2-Methylnaphthalene	0.707	0.704	0.4	234#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	231#	-0.01
14 S	2,4,6-Tribromophenol	0.042	0.064	-52.4#	487#	-0.01
15 S	2-Fluorobiphenyl	1.467	1.577	-7.5	239#	0.00
16	Acenaphthylene	1.617	1.567	3.1	229#	0.00
17	Acenaphthene	1.063	1.038	2.4	225#	0.00
18	Fluorene	1.453	1.350	7.1	214#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	210#	-0.01
20	4-Bromophenyl-phenylether	0.211	0.199	5.7	199#	0.00
21	Hexachlorobenzene	0.205	0.193	5.9	201#	0.00
22	Pentachlorophenol	0.000	0.000	0.0	0#	0.01
23	Phenanthrene	1.090	1.081	0.8	209#	0.00
24	Anthracene	0.974	0.919	5.6	207#	0.00
25 SURR	Fluoranthene-d10	5.287	4.502	14.8	182#	0.00
26	Fluoranthene	1.456	1.300	10.7	193#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	174#	0.00
28	Pvrene	1.112	1.278	-14.9	199#	0.00
29 S	Terphenyl-d14	0.954	1.066	-11.7	193#	0.00
30	Benzo(a)anthracene	1.284	1.237	3.7	166#	0.00
31	Chrysene	1.303	1.245	4.5	162#	0.00
32	Bis(2-ethylhexyl)phthalate	2.278	2.709	-18.9	229#	0.00
33	Indeno(1,2,3-cd)pyrene	1.527	1.456	4.6	161#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	174#	0.00
35	Benzo(b)fluoranthene	1.167	1.053	9.8	156#	0.00
36	Benzo(k)fluoranthene	1.223	1.154	5.6	164#	0.00
37 C	Benzo(a)pyrene	1.088	1.003	7.8	159#	0.00
38	Dibenzo(a,h)anthracene	1.134	1.052	7.2	161#	0.00
39	Benzo(a,h,i)perylene	1.113	1.044	6.2	161#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE111519\
Data File : BE100745.D
Acq On : 15 Nov 2019 20:00
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 15 23:53:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 15 15:55:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			