

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111618\
 Data File : BE098236.D
 Acq On : 16 Nov 2018 19:06
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE111618

Quant Time: Nov 16 19:38:11 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 16 19:05:33 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.628	0.643	-2.4	113	0.00
3	n-Nitrosodimethylamine	0.706	0.722	-2.3	115	-0.01
4 S	2-Fluorophenol	1.073	1.139	-6.2	110	0.00
5 S	Phenol-d6	1.685	1.762	-4.6	107	0.00
6	bis(2-Chloroethyl)ether	1.443	1.445	-0.1	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.378	0.410	-8.5	107	0.00
9	Naphthalene	4.073	4.244	-4.2	102	0.00
10	Hexachlorobutadiene	0.239	0.245	-2.5	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.702	0.721	-2.7	102	0.00
12	2-Methylnaphthalene	0.718	0.747	-4.0	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.129	0.137	-6.2	111	0.00
15 S	2-Fluorobiphenyl	1.739	1.738	0.1	103	0.00
16	Acenaphthylene	1.904	1.903	0.1	100	0.00
17	Acenaphthene	1.137	1.153	-1.4	99	0.00
18	Fluorene	1.449	1.484	-2.4	102	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
20	4-Bromophenyl-phenylether	0.248	0.251	-1.2	101	0.00
21	Hexachlorobenzene	0.256	0.262	-2.3	104	0.00
22	Pentachlorophenol	0.035	0.035	0.0	103	0.00
23	Phenanthrene	1.029	1.037	-0.8	101	0.00
24	Anthracene	0.935	0.943	-0.9	102	0.00
25 SURR	Fluoranthene-d10	4.866	4.920	-1.1	102	0.00
26	Fluoranthene	1.235	1.257	-1.8	102	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
28	Pvrene	1.169	1.205	-3.1	102	0.00
29 S	Terphenyl-d14	0.919	0.943	-2.6	103	0.00
30	Benzo(a)anthracene	1.157	1.181	-2.1	104	0.00
31	Chrysene	1.191	1.264	-6.1	106	0.00
32	Bis(2-ethylhexyl)phthalate	1.804	1.909	-5.8	112	0.00
33	Indeno(1,2,3-cd)pyrene	1.019	1.066	-4.6	105	0.00
34 I	Perylene-d12	1.000	1.000	0.0	105	0.00
35	Benzo(b)fluoranthene	1.130	1.138	-0.7	106	0.00
36	Benzo(k)fluoranthene	1.276	1.313	-2.9	105	0.00
37 C	Benzo(a)pyrene	1.109	1.113	-0.4	105	0.00
38	Dibenzo(a,h)anthracene	0.889	0.923	-3.8	110	0.00
39	Benzo(a,h,i)perylene	0.958	0.953	0.5	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111618\
Data File : BE098236.D
Acq On : 16 Nov 2018 19:06
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE111618

Quant Time: Nov 16 19:38:11 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 16 19:05:33 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			