

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111618\
 Data File : BE098242.D
 Acq On : 16 Nov 2018 22:50
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 16 23:20:43 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 16 19:05:33 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.628	0.584	7.0	106	0.00
3	n-Nitrosodimethylamine	0.706	0.740	-4.8	121	0.00
4 S	2-Fluorophenol	1.073	1.148	-7.0	115	0.00
5 S	Phenol-d6	1.685	1.895	-12.5	119	0.00
6	bis(2-Chloroethyl)ether	1.443	1.443	0.0	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.378	0.388	-2.6	108	0.00
9	Naphthalene	4.073	4.151	-1.9	106	0.00
10	Hexachlorobutadiene	0.239	0.242	-1.3	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.702	0.696	0.9	105	0.00
12	2-Methylnaphthalene	0.718	0.731	-1.8	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.129	0.168	-30.2#	139	-0.01
15 S	2-Fluorobiphenyl	1.739	1.683	3.2	102	0.00
16	Acenaphthylene	1.904	1.987	-4.4	106	0.00
17	Acenaphthene	1.137	1.216	-6.9	106	0.00
18	Fluorene	1.449	1.531	-5.7	108	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4-Bromophenyl-phenylether	0.248	0.256	-3.2	107	0.00
21	Hexachlorobenzene	0.256	0.264	-3.1	110	-0.01
22	Pentachlorophenol	0.035	0.058	-65.7#	178#	0.00
23	Phenanthrene	1.029	1.057	-2.7	107	-0.01
24	Anthracene	0.935	0.970	-3.7	109	0.00
25 SURR	Fluoranthene-d10	4.866	5.061	-4.0	109	0.00
26	Fluoranthene	1.235	1.257	-1.8	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
28	Pvrene	1.169	1.176	-0.6	108	0.00
29 S	Terphenyl-d14	0.919	0.935	-1.7	111	0.00
30	Benzo(a)anthracene	1.157	1.188	-2.7	114	0.00
31	Chrysene	1.191	1.223	-2.7	111	0.00
32	Bis(2-ethylhexyl)phthalate	1.804	2.593	-43.7#	166#	0.00
33	Indeno(1,2,3-cd)pyrene	1.019	1.156	-13.4	123	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	113	0.00
35	Benzo(b)fluoranthene	1.130	1.173	-3.8	119	0.00
36	Benzo(k)fluoranthene	1.276	1.253	1.8	109	0.00
37 C	Benzo(a)pyrene	1.109	1.139	-2.7	116	0.00
38	Dibenzo(a,h)anthracene	0.889	1.033	-16.2	134	-0.01
39	Benzo(a,h,i)perylene	0.958	1.015	-5.9	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111618\
Data File : BE098242.D
Acq On : 16 Nov 2018 22:50
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 16 23:20:43 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 16 19:05:33 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			