


Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111618\
 Data File : BE098228.D
 Acq On : 16 Nov 2018 13:53
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE111618.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Nov 16 19:05:33 2018



AutoFind: Scans 2577, 2578, 2579; Background Corrected with Scan 2569

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	53.5	12858	PASS
68	69	0.00	2	1.2	142	PASS
69	198	0.00	100	47.8	11489	PASS
70	69	0.00	2	0.7	77	PASS
127	198	10	80	45.7	10982	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	24041	PASS
199	198	5	9	6.9	1648	PASS
275	198	10	60	30.2	7268	PASS
365	198	1	100	4.3	1033	PASS
441	198	0.01	100	13.8	3312	PASS
442	198	50	100	86.3	20755	PASS
443	442	15	24	18.7	3880	PASS