

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE112219\
 Data File : BE100768.D
 Acq On : 22 Nov 2019 21:26
 Operator : JU
 Sample : K5969-08
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 HF-104S

Quant Time: Nov 23 04:48:12 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 22 06:40:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.76	152	6802	0.40	ng	0.00
7) Naphthalene-d8	10.55	136	31583	0.40	ng	0.00
13) Acenaphthene-d10	14.40	164	20343	0.40	ng	0.00
19) Phenanthrene-d10	17.13	188	42375	0.40	ng	0.00
27) Chrysene-d12	21.32	240	39712	0.40	ng	0.00
34) Perylene-d12	23.77	264	45766	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.38	112	3297	0.18	ng	0.00
5) Phenol-d6	6.93	99	2875	0.11	ng	0.00
8) Nitrobenzene-d5	8.90	82	8575	0.85	ng	0.00
11) 2-Methylnaphthalene-d10	12.14	152	17859	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.88	330	2911	1.02	ng	0.00
15) 2-Fluorobiphenyl	13.03	172	25105	0.32	ng	0.00
25) Fluoranthene-d10	19.17	212	200406	0.39	ng	0.00
29) Terphenyl-d14	19.77	244	40320	0.44	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.34	88	63392	4.968	ng	98
9) Naphthalene	10.59	128	17198	0.051	ng	# 95
12) 2-Methylnaphthalene	12.21	142	1475	0.028	ng	# 69
17) Acenaphthene	14.47	154	32318	0.550	ng	98
24) Anthracene	17.27	178	5789	0.053	ng	# 80
26) Fluoranthene	19.20	202	27291	0.175	ng	# 95
28) Pyrene	19.56	202	19627	0.160	ng	# 89
30) Benzo(a)anthracene	21.30	228	2968	0.022	ng	# 44
32) Bis(2-ethylhexyl)phthalate	21.26	149	29230	0.123	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE112219\
Data File : BE100768.D
Acq On : 22 Nov 2019 21:26
Operator : JU
Sample : K5969-08
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
HF-104S

Quant Time: Nov 23 04:48:12 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 22 06:40:44 2019
Response via : Initial Calibration

