

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE112219\
 Data File : BE100756.D
 Acq On : 22 Nov 2019 13:14
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 22 14:42:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 22 06:40:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.750	0.821	-9.5	118	0.00
3	n-Nitrosodimethylamine	0.553	0.512	7.4	117	0.00
4 S	2-Fluorophenol	1.083	1.033	4.6	105	0.00
5 S	Phenol-d6	1.571	1.506	4.1	107	0.00
6	bis(2-Chloroethyl)ether	1.310	1.277	2.5	114	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.128	0.168	-31.3#	165#	0.00
9	Naphthalene	4.248	4.247	0.0	107	0.00
10	Hexachlorobutadiene	0.163	0.168	-3.1	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.587	0.565	3.7	108	0.00
12	2-Methylnaphthalene	0.672	0.656	2.4	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.056	0.058	-3.6	125	0.00
15 S	2-Fluorobiphenyl	1.553	1.602	-3.2	112	0.00
16	Acenaphthylene	1.722	1.519	11.8	102	0.00
17	Acenaphthene	1.155	1.112	3.7	107	0.00
18	Fluorene	1.475	1.389	5.8	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4-Bromophenyl-phenylether	0.197	0.191	3.0	106	0.01
21	Hexachlorobenzene	0.205	0.205	0.0	108	0.00
22	Pentachlorophenol	0.037	0.033	10.8	128	0.00
23	Phenanthrene	1.205	1.198	0.6	108	0.00
24	Anthracene	1.038	0.957	7.8	107	0.00
25 SURR	Fluoranthene-d10	4.896	4.526	7.6	103	0.00
26	Fluoranthene	1.474	1.370	7.1	105	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
28	Pvrene	1.239	1.230	0.7	102	0.00
29 S	Terphenyl-d14	0.921	0.918	0.3	102	0.00
30	Benzo(a)anthracene	1.349	1.277	5.3	96	-0.01
31	Chrysene	1.372	1.373	-0.1	100	0.00
32	Bis(2-ethylhexyl)phthalate	2.427	1.220	49.7#	72	0.00
33	Indeno(1,2,3-cd)pyrene	1.438	1.113	22.6	90	0.00
34 I	Perylene-d12	1.000	1.000	0.0	96	0.00
35	Benzo(b)fluoranthene	1.255	1.180	6.0	91	0.00
36	Benzo(k)fluoranthene	1.309	1.260	3.7	95	0.00
37 C	Benzo(a)pyrene	1.107	0.985	11.0	89	0.00
38	Dibenzo(a,h)anthracene	1.164	0.977	16.1	90	0.00
39	Benzo(a,h,i)perylene	1.213	1.060	12.6	93	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE112219\
Data File : BE100756.D
Acq On : 22 Nov 2019 13:14
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 22 14:42:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 22 06:40:44 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			