

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE112219\
 Data File : BE100769.D
 Acq On : 22 Nov 2019 22:41
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 23 04:26:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 22 06:40:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.750	0.743	0.9	105	0.00
3	n-Nitrosodimethylamine	0.553	0.624	-12.8	140	0.00
4 S	2-Fluorophenol	1.083	1.209	-11.6	121	0.00
5 S	Phenol-d6	1.571	1.744	-11.0	122	0.00
6	bis(2-Chloroethyl)ether	1.310	1.377	-5.1	120	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
8 S	Nitrobenzene-d5	0.128	0.246	-92.2#	245#	0.00
9	Naphthalene	4.248	4.349	-2.4	111	0.00
10	Hexachlorobutadiene	0.163	0.171	-4.9	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.587	0.600	-2.2	116	0.00
12	2-Methylnaphthalene	0.672	0.671	0.1	112	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.056	0.078	-39.3#	178#	0.00
15 S	2-Fluorobiphenyl	1.553	1.567	-0.9	115	0.00
16	Acenaphthylene	1.722	1.660	3.6	117	0.00
17	Acenaphthene	1.155	1.134	1.8	115	-0.02
18	Fluorene	1.475	1.437	2.6	116	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
20	4-Bromophenyl-phenylether	0.197	0.188	4.6	113	0.00
21	Hexachlorobenzene	0.205	0.200	2.4	114	0.00
22	Pentachlorophenol	0.037	0.019	48.6#	79	0.00
23	Phenanthrene	1.205	1.205	0.0	118	0.00
24	Anthracene	1.038	1.005	3.2	121	0.00
25 SURR	Fluoranthene-d10	4.896	4.705	3.9	117	0.00
26	Fluoranthene	1.474	1.399	5.1	116	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
28	Pvrene	1.239	1.375	-11.0	122	0.00
29 S	Terphenyl-d14	0.921	0.959	-4.1	113	0.00
30	Benzo(a)anthracene	1.349	1.347	0.1	108	-0.01
31	Chrysene	1.372	1.375	-0.2	106	0.00
32	Bis(2-ethylhexyl)phthalate	2.427	1.866	23.1	117	0.00
33	Indeno(1,2,3-cd)pyrene	1.438	1.280	11.0	111	0.00
34 I	Perylene-d12	1.000	1.000	0.0	122	0.00
35	Benzo(b)fluoranthene	1.255	1.095	12.7	107	0.00
36	Benzo(k)fluoranthene	1.309	1.144	12.6	109	0.00
37 C	Benzo(a)pyrene	1.107	0.962	13.1	110	0.00
38	Dibenzo(a,h)anthracene	1.164	0.941	19.2	109	0.00
39	Benzo(a,h,i)perylene	1.213	0.998	17.7	110	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE112219\
Data File : BE100769.D
Acq On : 22 Nov 2019 22:41
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 23 04:26:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 22 06:40:44 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			