

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE112918\
 Data File : BE098274.D
 Acq On : 29 Nov 2018 15:12
 Operator : SJ/JU
 Sample : PB115083BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB115083BL

Quant Time: Nov 29 16:17:31 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 29 13:50:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.86	152	1420	0.40	ng	0.00
7) Naphthalene-d8	10.67	136	5363	0.40	ng	0.00
13) Acenaphthene-d10	14.53	164	3006	0.40	ng	0.00
19) Phenanthrene-d10	17.28	188	7028	0.40	ng	0.00
27) Chrysene-d12	21.47	240	9913	0.40	ng	0.01
34) Perylene-d12	24.02	264	10221	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.46	112	1052	0.28	ng	0.01
5) Phenol-d6	7.06	99	1240	0.23	ng	0.01
8) Nitrobenzene-d5	9.03	82	1440	0.27	ng	0.01
11) 2-Methylnaphthalene-d10	12.27	152	2567	0.29	ng	0.00
14) 2,4,6-Tribromophenol	16.03	330	257	0.18	ng	0.01
15) 2-Fluorobiphenyl	13.15	172	3637	0.29	ng	0.00
25) Fluoranthene-d10	19.31	212	29236	0.31	ng	0.00
29) Terphenyl-d14	19.91	244	5127	0.23	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE112918\
Data File : BE098274.D
Acq On : 29 Nov 2018 15:12
Operator : SJ/JU
Sample : PB115083BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB115083BL

Quant Time: Nov 29 16:17:31 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 29 13:50:43 2018
Response via : Initial Calibration

