

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE112918\
 Data File : BE098272.D
 Acq On : 29 Nov 2018 13:58
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE112918

Quant Time: Nov 29 15:06:12 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 29 13:50:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	81	0.00
2	1,4-Dioxane	0.400	0.410	-2.5	87	0.00
3	n-Nitrosodimethylamine	0.400	0.364	9.0	75	0.00
4 S	2-Fluorophenol	0.400	0.388	3.0	79	0.00
5 S	Phenol-d6	0.400	0.405	-1.3	85	0.00
6	bis(2-Chloroethyl)ether	0.400	0.376	6.0	77	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	80	0.00
8 S	Nitrobenzene-d5	0.400	0.404	-1.0	83	0.00
9	Naphthalene	0.400	0.402	-0.5	81	0.00
10	Hexachlorobutadiene	0.400	0.394	1.5	80	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	75	0.00
12	2-Methylnaphthalene	0.400	0.377	5.8	76	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	75	0.00
14 S	2,4,6-Tribromophenol	0.400	0.369	7.8	79	0.00
15 S	2-Fluorobiphenyl	0.400	0.391	2.3	79	0.00
16	Acenaphthylene	0.400	0.389	2.8	75	0.00
17	Acenaphthene	0.400	0.399	0.3	77	0.01
18	Fluorene	0.400	0.389	2.8	74	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	76	0.00
20	4-Bromophenyl-phenylether	0.400	0.381	4.8	73	0.00
21	Hexachlorobenzene	0.400	0.395	1.3	75	0.00
22	Pentachlorophenol	0.400	0.393	1.8	89	0.00
23	Phenanthrene	0.400	0.392	2.0	75	0.01
24	Anthracene	0.400	0.382	4.5	74	0.01
25 SURR	Fluoranthene-d10	0.400	0.404	-1.0	78	0.00
26	Fluoranthene	0.400	0.401	-0.3	79	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	94	0.00
28	Pvrene	0.400	0.368	8.0	81	0.00
29 S	Terphenyl-d14	0.400	0.353	11.8	80	0.00
30	Benzo(a)anthracene	0.400	0.380	5.0	92	0.00
31	Chrysene	0.400	0.389	2.8	93	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.384	4.0	92	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.411	-2.7	104	0.00
34 I	Perylene-d12	0.400	0.400	0.0	102	0.00
35	Benzo(b)fluoranthene	0.400	0.380	5.0	99	0.00
36	Benzo(k)fluoranthene	0.400	0.397	0.8	105	0.00
37 C	Benzo(a)pyrene	0.400	0.391	2.3	102	0.00
38	Dibenzo(a,h)anthracene	0.400	0.393	1.8	103	0.00
39	Benzo(a,h,i)perylene	0.400	0.390	2.5	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE112918\
Data File : BE098272.D
Acq On : 29 Nov 2018 13:58
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE112918

Quant Time: Nov 29 15:06:12 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 29 13:50:43 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		