

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE112918\
 Data File : BE098281.D
 Acq On : 29 Nov 2018 19:32
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 30 00:30:43 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 29 13:50:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	0.742	0.589	20.6	112	0.00
3	n-Nitrosodimethylamine	0.768	0.688	10.4	116	0.00
4 S	2-Fluorophenol	1.068	1.059	0.8	128	0.01
5 S	Phenol-d6	1.505	1.539	-2.3	135	0.00
6	bis(2-Chloroethyl)ether	1.385	1.333	3.8	125	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
8 S	Nitrobenzene-d5	0.405	0.391	3.5	127	0.00
9	Naphthalene	4.019	3.993	0.6	128	-0.01
10	Hexachlorobutadiene	0.258	0.250	3.1	126	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.664	-2.2	127	0.00
12	2-Methylnaphthalene	0.667	0.660	1.0	127	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	130	0.00
14 S	2,4,6-Tribromophenol	0.190	0.174	8.4	136	0.00
15 S	2-Fluorobiphenyl	1.657	1.556	6.1	131	0.00
16	Acenaphthylene	1.867	1.807	3.2	129	-0.01
17	Acenaphthene	1.113	1.082	2.8	130	0.00
18	Fluorene	1.475	1.474	0.1	131	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
20	4-Bromophenyl-phenylether	0.226	0.219	3.1	127	0.00
21	Hexachlorobenzene	0.234	0.228	2.6	125	0.00
22	Pentachlorophenol	0.076	0.065	14.5	141	0.00
23	Phenanthrene	0.993	0.972	2.1	128	0.00
24	Anthracene	0.925	0.900	2.7	129	0.00
25 SURR	Fluoranthene-d10	5.397	5.015	7.1	123	0.00
26	Fluoranthene	1.356	1.250	7.8	123	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
28	Pvrene	1.108	1.227	-10.7	123	0.00
29 S	Terphenyl-d14	0.900	0.943	-4.8	121	0.00
30	Benzo(a)anthracene	1.178	1.127	4.3	118	0.00
31	Chrysene	1.196	1.154	3.5	118	-0.01
32	Bis(2-ethylhexyl)phthalate	2.821	2.793	1.0	121	0.00
33	Indeno(1,2,3-cd)pyrene	1.210	1.138	6.0	121	0.00
34 I	Perylene-d12	1.000	1.000	0.0	120	0.00
35	Benzo(b)fluoranthene	1.173	1.147	2.2	119	0.00
36	Benzo(k)fluoranthene	1.231	1.213	1.5	122	0.00
37 C	Benzo(a)pyrene	1.130	1.118	1.1	121	0.00
38	Dibenzo(a,h)anthracene	1.071	1.049	2.1	120	0.00
39	Benzo(a,h,i)perylene	1.073	1.046	2.5	120	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE112918\
Data File : BE098281.D
Acq On : 29 Nov 2018 19:32
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 30 00:30:43 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 29 13:50:43 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			