

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE120519\
 Data File : BE100788.D
 Acq On : 5 Dec 2019 14:34
 Operator : JU
 Sample : K6024-08
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 TWP-6

Manual Integrations
 APPROVED

mohammad
 12/6/2019 3:02:30 PM

Quant Time: Dec 06 10:01:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.76	152	6361	0.40	ng	-0.01
7) Naphthalene-d8	10.54	136	30125	0.40	ng	0.00
13) Acenaphthene-d10	14.40	164	20739	0.40	ng	0.00
19) Phenanthrene-d10	17.12	188	49503	0.40	ng	-0.01
27) Chrysene-d12	21.31	240	46178	0.40	ng	0.00
34) Perylene-d12	23.76	264	52417	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.37	112	4141	0.27	ng	0.00
5) Phenol-d6	6.93	99	4456	0.20	ng	0.00
8) Nitrobenzene-d5	8.90	82	6506	0.45	ng	0.00
11) 2-Methylnaphthalene-d10	12.14	152	16381	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.88	330	3286	1.42	ng	0.00
15) 2-Fluorobiphenyl	13.02	172	24249	0.29	ng	-0.01
25) Fluoranthene-d10	19.16	212	217803	0.37	ng	0.00
29) Terphenyl-d14	19.76	244	45182	0.41	ng	0.00
Target Compounds						
9) Naphthalene	10.59	128	9294	0.029	ng	# 95
12) 2-Methylnaphthalene	12.21	142	1648	0.033	ng	# 78
22) Pentachlorophenol	16.79	266	182m	0.100	ng	
24) Anthracene	17.25	178	4632	0.037	ng	# 39
32) Bis(2-ethylhexyl)phthalate	21.25	149	41386	0.433	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE120519\
Data File : BE100788.D
Acq On : 5 Dec 2019 14:34
Operator : JU
Sample : K6024-08
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
TWP-6

Manual Integrations
APPROVED

mohammad
12/6/2019 3:02:30 PM

Quant Time: Dec 06 10:01:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

