




Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE120519\
Data File : BE100780.D
Acq On : 5 Dec 2019 8:58
Operator : JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 05 17:11:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration



(22) Pentachlorophenol

16.789min (+0.002) 54141.90ng

response 69862

Ion	Exp%	Act%
266.00	100	100
264.00	79.20	63.26#
268.00	82.30	69.89
0.00	0.00	0.00