

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE120519\
 Data File : BE100781.D
 Acq On : 5 Dec 2019 10:14
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 05 12:50:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.729	0.680	6.7	94	0.00
3	n-Nitrosodimethylamine	0.594	0.525	11.6	102	0.00
4 S	2-Fluorophenol	0.949	0.950	-0.1	107	0.00
5 S	Phenol-d6	1.433	1.423	0.7	108	0.00
6	bis(2-Chloroethyl)ether	1.385	1.366	1.4	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.190	0.187	1.6	120	0.00
9	Naphthalene	4.250	4.058	4.5	106	0.00
10	Hexachlorobutadiene	0.166	0.156	6.0	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.547	5.7	106	0.00
12	2-Methylnaphthalene	0.664	0.611	8.0	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.044	0.038	13.6	105	0.00
15 S	2-Fluorobiphenyl	1.587	1.488	6.2	105	0.00
16	Acenaphthylene	1.655	1.425	13.9	107	0.00
17	Acenaphthene	1.140	1.057	7.3	109	0.00
18	Fluorene	1.471	1.340	8.9	109	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	108	-0.01
20	4-Bromophenyl-phenylether	0.196	0.185	5.6	113	0.00
21	Hexachlorobenzene	0.213	0.203	4.7	109	0.00
22	Pentachlorophenol	0.015	0.003	80.0#	132	0.00
23	Phenanthrene	1.212	1.148	5.3	109	0.00
24	Anthracene	1.013	0.881	13.0	110	0.00
25 SURR	Fluoranthene-d10	4.728	4.184	11.5	106	0.00
26	Fluoranthene	1.446	1.288	10.9	107	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
28	Pvrene	1.300	1.232	5.2	107	0.00
29 S	Terphenyl-d14	0.958	0.921	3.9	106	0.00
30	Benzo(a)anthracene	1.179	1.075	8.8	108	0.00
31	Chrysene	1.385	1.334	3.7	105	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	0.886	17.1	115	0.00
33	Indeno(1,2,3-cd)pyrene	1.207	1.087	9.9	108	0.00
34 I	Perylene-d12	1.000	1.000	0.0	109	0.00
35	Benzo(b)fluoranthene	1.095	1.016	7.2	107	0.00
36	Benzo(k)fluoranthene	1.228	1.138	7.3	106	0.00
37 C	Benzo(a)pyrene	0.948	0.850	10.3	107	0.00
38	Dibenzo(a,h)anthracene	0.943	0.822	12.8	107	-0.02
39	Benzo(a,h,i)perylene	1.009	0.950	5.8	109	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE120519\
Data File : BE100781.D
Acq On : 5 Dec 2019 10:14
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 05 12:50:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			