

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE120519\
 Data File : BE100797.D
 Acq On : 5 Dec 2019 20:08
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 06 02:05:09 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	111	-0.01
2	1,4-Dioxane	0.400	0.334	16.5	93	-0.01
3	n-Nitrosodimethylamine	0.400	0.378	5.5	120	-0.01
4 S	2-Fluorophenol	0.400	0.513	-28.2#	152	0.00
5 S	Phenol-d6	0.400	0.518	-29.5#	155	0.00
6	bis(2-Chloroethyl)ether	0.400	0.407	-1.7	119	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	122	0.00
8 S	Nitrobenzene-d5	0.400	0.624	-56.0#	219	0.00
9	Naphthalene	0.400	0.380	5.0	121	0.00
10	Hexachlorobutadiene	0.400	0.363	9.3	114	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.402	-0.5	131	0.00
12	2-Methylnaphthalene	0.400	0.391	2.3	127	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	140	0.00
14 S	2,4,6-Tribromophenol	0.400	0.957	-139.2#	383	0.00
15 S	2-Fluorobiphenyl	0.400	0.343	14.2	125	0.00
16	Acenaphthylene	0.400	0.396	1.0	161	0.00
17	Acenaphthene	0.400	0.372	7.0	142	0.00
18	Fluorene	0.400	0.373	6.8	146	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	128	-0.01
20	4-Bromophenyl-phenylether	0.400	0.393	1.8	139	0.00
21	Hexachlorobenzene	0.400	0.376	6.0	128	0.00
22	Pentachlorophenol	0.400	1.084	-171.0#	1796	0.00
23	Phenanthrene	0.400	0.378	5.5	129	0.00
24	Anthracene	0.400	0.391	2.3	146	0.00
25 SURR	Fluoranthene-d10	0.400	0.373	6.8	132	0.00
26	Fluoranthene	0.400	0.356	11.0	127	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	117	0.00
28	Pvrene	0.400	0.424	-6.0	133	0.00
29 S	Terphenyl-d14	0.400	0.397	0.8	121	0.00
30	Benzo(a)anthracene	0.400	0.461	-15.3	152	0.00
31	Chrysene	0.400	0.390	2.5	118	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	1.322	-230.5#	467	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.503	-25.7#	167	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	140	0.00
35	Benzo(b)fluoranthene	0.400	0.397	0.8	146	0.00
36	Benzo(k)fluoranthene	0.400	0.365	8.8	134	0.00
37 C	Benzo(a)pyrene	0.400	0.430	-7.5	165	0.00
38	Dibenzo(a,h)anthracene	0.400	0.429	-7.2	169	-0.01
39	Benzo(a,h,i)perylene	0.400	0.407	-1.7	151	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE120519\
Data File : BE100797.D
Acq On : 5 Dec 2019 20:08
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 06 02:05:09 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		