

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE120519\
 Data File : BE100797.D
 Acq On : 5 Dec 2019 20:08
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 06 02:05:09 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	-0.01
2	1,4-Dioxane	0.729	0.609	16.5	93	-0.01
3	n-Nitrosodimethylamine	0.594	0.561	5.6	120	-0.01
4 S	2-Fluorophenol	0.949	1.218	-28.3#	152#	0.00
5 S	Phenol-d6	1.433	1.855	-29.4#	155#	0.00
6	bis(2-Chloroethyl)ether	1.385	1.410	-1.8	119	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
8 S	Nitrobenzene-d5	0.190	0.297	-56.3#	219#	0.00
9	Naphthalene	4.250	4.037	5.0	121	0.00
10	Hexachlorobutadiene	0.166	0.150	9.6	114	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.582	-0.3	131	0.00
12	2-Methylnaphthalene	0.664	0.649	2.3	127	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	140	0.00
14 S	2,4,6-Tribromophenol	0.044	0.106	-140.9#	383#	0.00
15 S	2-Fluorobiphenyl	1.587	1.360	14.3	125	0.00
16	Acenaphthylene	1.655	1.640	0.9	161#	0.00
17	Acenaphthene	1.140	1.060	7.0	142	0.00
18	Fluorene	1.471	1.371	6.8	146	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	128	-0.01
20	4-Bromophenyl-phenylether	0.196	0.193	1.5	139	0.00
21	Hexachlorobenzene	0.213	0.200	6.1	128	0.00
22	Pentachlorophenol	0.015	0.040	-166.7#	1796#	0.00
23	Phenanthrene	1.212	1.145	5.5	129	0.00
24	Anthracene	1.013	0.991	2.2	146	0.00
25 SURR	Fluoranthene-d10	4.728	4.403	6.9	132	0.00
26	Fluoranthene	1.446	1.287	11.0	127	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
28	Pvrene	1.300	1.379	-6.1	133	0.00
29 S	Terphenyl-d14	0.958	0.951	0.7	121	0.00
30	Benzo(a)anthracene	1.179	1.360	-15.4	152#	0.00
31	Chrysene	1.385	1.352	2.4	118	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	3.237	-202.8#	467#	0.00
33	Indeno(1,2,3-cd)pyrene	1.207	1.517	-25.7#	167#	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	140	0.00
35	Benzo(b)fluoranthene	1.095	1.087	0.7	146	0.00
36	Benzo(k)fluoranthene	1.228	1.122	8.6	134	0.00
37 C	Benzo(a)pyrene	0.948	1.019	-7.5	165#	0.00
38	Dibenzo(a,h)anthracene	0.943	1.011	-7.2	169#	-0.01
39	Benzo(a,h,i)perylene	1.009	1.027	-1.8	151#	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE120519\
Data File : BE100797.D
Acq On : 5 Dec 2019 20:08
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 06 02:05:09 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			