

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121018\
 Data File : BE098303.D
 Acq On : 10 Dec 2018 18:11
 Operator : SJ/JU
 Sample : J6259-10
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BPS1-TT-MW306D-20181205

Quant Time: Dec 11 01:04:53 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.86	152	1495	0.40	ng	0.00
7) Naphthalene-d8	10.67	136	5937	0.40	ng	0.00
13) Acenaphthene-d10	14.53	164	3894	0.40	ng	0.00
19) Phenanthrene-d10	17.28	188	10049	0.40	ng	0.00
27) Chrysene-d12	21.46	240	12461	0.40	ng	0.00
34) Perylene-d12	24.01	264	11949	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.46	112	483	0.14	ng	0.01
5) Phenol-d6	7.05	99	390	0.08	ng	0.00
8) Nitrobenzene-d5	9.03	82	1882	0.35	ng	0.01
11) 2-Methylnaphthalene-d10	12.27	152	3379	0.35	ng	0.00
14) 2,4,6-Tribromophenol	16.02	330	502	0.35	ng	0.00
15) 2-Fluorobiphenyl	13.15	172	7014	0.43	ng	0.00
25) Fluoranthene-d10	19.31	212	49711	0.39	ng	0.00
29) Terphenyl-d14	19.90	244	11943	0.39	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.37	88	2406	0.931	ng	# 86
26) Fluoranthene	19.34	202	656	0.020	ng	# 92
30) Benzo(a)anthracene	21.45	228	777	0.022	ng	# 83
32) Bis(2-ethylhexyl)phthalate	21.37	149	21865	0.303	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121018\
Data File : BE098303.D
Acq On : 10 Dec 2018 18:11
Operator : SJ/JU
Sample : J6259-10
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BPS1-TT-MW306D-20181205

Quant Time: Dec 11 01:04:53 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

