

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121018\
 Data File : BE098301.D
 Acq On : 10 Dec 2018 16:54
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 11 00:47:35 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.692	0.660	4.6	101	0.00
3	n-Nitrosodimethylamine	0.623	0.622	0.2	112	0.00
4 S	2-Fluorophenol	0.936	0.933	0.3	106	0.00
5 S	Phenol-d6	1.326	1.294	2.4	100	0.00
6	bis(2-Chloroethyl)ether	1.260	1.205	4.4	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.364	0.359	1.4	112	0.01
9	Naphthalene	4.025	4.039	-0.3	107	0.00
10	Hexachlorobutadiene	0.256	0.261	-2.0	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.644	0.9	105	0.00
12	2-Methylnaphthalene	0.654	0.650	0.6	106	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.147	0.138	6.1	103	0.00
15 S	2-Fluorobiphenyl	1.687	1.583	6.2	107	0.00
16	Acenaphthylene	1.874	1.837	2.0	106	0.00
17	Acenaphthene	1.126	1.127	-0.1	105	0.00
18	Fluorene	1.506	1.485	1.4	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4-Bromophenyl-phenylether	0.215	0.207	3.7	102	0.00
21	Hexachlorobenzene	0.231	0.233	-0.9	104	0.00
22	Pentachlorophenol	0.046	0.033	28.3#	93	0.00
23	Phenanthrene	0.972	0.947	2.6	100	0.00
24	Anthracene	0.866	0.820	5.3	97	0.00
25 SURR	Fluoranthene-d10	5.131	5.092	0.8	97	0.00
26	Fluoranthene	1.286	1.274	0.9	96	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
28	Pvrene	1.255	1.289	-2.7	97	0.00
29 S	Terphenyl-d14	0.972	0.934	3.9	93	0.00
30	Benzo(a)anthracene	1.139	1.090	4.3	94	0.00
31	Chrysene	1.193	1.198	-0.4	98	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	2.137	7.6	90	0.00
33	Indeno(1,2,3-cd)pyrene	1.187	1.192	-0.4	102	0.00
34 I	Perylene-d12	1.000	1.000	0.0	101	0.00
35	Benzo(b)fluoranthene	1.094	1.070	2.2	102	0.00
36	Benzo(k)fluoranthene	1.274	1.245	2.3	103	0.00
37 C	Benzo(a)pyrene	1.129	1.109	1.8	101	0.00
38	Dibenzo(a,h)anthracene	1.063	1.033	2.8	101	0.00
39	Benzo(a,h,i)perylene	1.071	1.063	0.7	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121018\
Data File : BE098301.D
Acq On : 10 Dec 2018 16:54
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 11 00:47:35 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			