

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121018\
 Data File : BE098318.D
 Acq On : 11 Dec 2018 4:02
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 11 07:23:57 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.692	0.676	2.3	105	0.00
3	n-Nitrosodimethylamine	0.623	0.494	20.7	90	0.00
4 S	2-Fluorophenol	0.936	0.960	-2.6	112	0.00
5 S	Phenol-d6	1.326	1.322	0.3	105	0.00
6	bis(2-Chloroethyl)ether	1.260	1.123	10.9	98	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.364	0.357	1.9	109	0.01
9	Naphthalene	4.025	3.954	1.8	103	0.00
10	Hexachlorobutadiene	0.256	0.254	0.8	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.643	1.1	102	0.00
12	2-Methylnaphthalene	0.654	0.616	5.8	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.147	0.142	3.4	98	0.00
15 S	2-Fluorobiphenyl	1.687	1.584	6.1	99	0.00
16	Acenaphthylene	1.874	1.799	4.0	95	0.00
17	Acenaphthene	1.126	1.055	6.3	91	0.00
18	Fluorene	1.506	1.429	5.1	91	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4-Bromophenyl-phenylether	0.215	0.208	3.3	89	0.00
21	Hexachlorobenzene	0.231	0.235	-1.7	91	0.00
22	Pentachlorophenol	0.046	0.041	10.9	101	0.00
23	Phenanthrene	0.972	0.938	3.5	86	0.00
24	Anthracene	0.866	0.823	5.0	85	0.00
25 SURR	Fluoranthene-d10	5.131	5.393	-5.1	89	0.00
26	Fluoranthene	1.286	1.362	-5.9	90	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
28	Pyrene	1.255	1.096	12.7	89	0.00
29 S	Terphenyl-d14	0.972	0.817	15.9	89	0.00
30	Benzo(a)anthracene	1.139	1.067	6.3	100	0.00
31	Chrysene	1.193	1.169	2.0	103	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	2.214	4.3	101	0.00
33	Indeno(1,2,3-cd)pyrene	1.187	1.187	0.0	110	0.00
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.094	1.064	2.7	112	0.00
36	Benzo(k)fluoranthene	1.274	1.281	-0.5	117	0.00
37 C	Benzo(a)pyrene	1.129	1.076	4.7	109	0.00
38	Dibenzo(a,h)anthracene	1.063	1.025	3.6	111	0.00
39	Benzo(a,h,i)perylene	1.071	1.033	3.5	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121018\
Data File : BE098318.D
Acq On : 11 Dec 2018 4:02
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 11 07:23:57 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		