

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121018\
 Data File : BE098332.D
 Acq On : 11 Dec 2018 14:34
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 11 15:23:23 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	102	0.00
2	1,4-Dioxane	0.400	0.425	-6.2	110	0.00
3	n-Nitrosodimethylamine	0.400	0.296	26.0#	81	0.00
4 S	2-Fluorophenol	0.400	0.372	7.0	98	0.00
5 S	Phenol-d6	0.400	0.370	7.5	94	0.00
6	bis(2-Chloroethyl)ether	0.400	0.348	13.0	92	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	0.00
8 S	Nitrobenzene-d5	0.400	0.367	8.3	101	0.01
9	Naphthalene	0.400	0.392	2.0	101	0.01
10	Hexachlorobutadiene	0.400	0.402	-0.5	104	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.387	3.3	99	0.00
12	2-Methylnaphthalene	0.400	0.367	8.3	95	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.400	0.369	7.8	94	0.01
15 S	2-Fluorobiphenyl	0.400	0.377	5.8	99	0.00
16	Acenaphthylene	0.400	0.389	2.8	97	0.02
17	Acenaphthene	0.400	0.389	2.8	94	0.00
18	Fluorene	0.400	0.386	3.5	93	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	94	0.00
20	4-Bromophenyl-phenylether	0.400	0.379	5.3	93	0.01
21	Hexachlorobenzene	0.400	0.388	3.0	93	0.00
22	Pentachlorophenol	0.400	0.355	11.3	84	0.00
23	Phenanthrene	0.400	0.391	2.3	93	0.00
24	Anthracene	0.400	0.367	8.3	87	0.01
25 SURR	Fluoranthene-d10	0.400	0.391	2.3	88	0.00
26	Fluoranthene	0.400	0.397	0.8	89	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	106	0.00
28	Pvrene	0.400	0.353	11.8	91	0.00
29 S	Terphenyl-d14	0.400	0.336	16.0	90	0.00
30	Benzo(a)anthracene	0.400	0.371	7.3	100	0.00
31	Chrysene	0.400	0.404	-1.0	108	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.342	14.5	91	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.418	-4.5	117	0.00
34 I	Perylene-d12	0.400	0.400	0.0	115	0.00
35	Benzo(b)fluoranthene	0.400	0.395	1.3	117	0.00
36	Benzo(k)fluoranthene	0.400	0.418	-4.5	125	0.00
37 C	Benzo(a)pyrene	0.400	0.397	0.8	117	0.00
38	Dibenzo(a,h)anthracene	0.400	0.394	1.5	117	0.00
39	Benzo(a,h,i)perylene	0.400	0.394	1.5	117	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121018\
Data File : BE098332.D
Acq On : 11 Dec 2018 14:34
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 11 15:23:23 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			