


Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121018\
 Data File : BE098317.D
 Acq On : 11 Dec 2018 3:26
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121018.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Dec 10 14:59:55 2018



AutoFind: Scans 2575, 2576, 2577; Background Corrected with Scan 2566

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	53.1	24223	PASS
68	69	0.00	2	0.7	146	PASS
69	198	0.00	100	48.1	21945	PASS
70	69	0.00	2	0.4	78	PASS
127	198	10	80	45.7	20854	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	45641	PASS
199	198	5	9	6.9	3133	PASS
275	198	10	60	29.9	13641	PASS
365	198	1	100	4.5	2040	PASS
441	198	0.01	100	15.0	6861	PASS
442	198	50	100	90.7	41404	PASS
443	442	15	24	19.2	7944	PASS