

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121218\
 Data File : BE098362.D
 Acq On : 12 Dec 2018 15:42
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 13 05:51:55 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	81	0.00
2	1,4-Dioxane	0.400	0.415	-3.7	85	0.00
3	n-Nitrosodimethylamine	0.400	0.269	32.8#	58	0.00
4 S	2-Fluorophenol	0.400	0.383	4.3	80	0.01
5 S	Phenol-d6	0.400	0.363	9.3	73	0.00
6	bis(2-Chloroethyl)ether	0.400	0.334	16.5	70	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	78	0.00
8 S	Nitrobenzene-d5	0.400	0.382	4.5	82	0.01
9	Naphthalene	0.400	0.383	4.3	77	0.00
10	Hexachlorobutadiene	0.400	0.392	2.0	79	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.379	5.3	75	0.00
12	2-Methylnaphthalene	0.400	0.363	9.3	73	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	74	0.00
14 S	2,4,6-Tribromophenol	0.400	0.340	15.0	66	0.00
15 S	2-Fluorobiphenyl	0.400	0.363	9.3	74	0.00
16	Acenaphthylene	0.400	0.384	4.0	73	0.00
17	Acenaphthene	0.400	0.374	6.5	70	0.00
18	Fluorene	0.400	0.382	4.5	71	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	71	0.00
20	4-Bromophenyl-phenylether	0.400	0.360	10.0	67	0.00
21	Hexachlorobenzene	0.400	0.384	4.0	70	0.00
22	Pentachlorophenol	0.400	0.363	9.3	65	0.01
23	Phenanthrene	0.400	0.370	7.5	67	0.00
24	Anthracene	0.400	0.351	12.3	63	0.00
25 SURR	Fluoranthene-d10	0.400	0.383	4.3	65	0.00
26	Fluoranthene	0.400	0.387	3.3	66	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	81	0.00
28	Pvrene	0.400	0.338	15.5	67	0.00
29 S	Terphenyl-d14	0.400	0.316	21.0	65	0.00
30	Benzo(a)anthracene	0.400	0.353	11.8	73	0.00
31	Chrysene	0.400	0.382	4.5	79	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.340	15.0	70	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.382	4.5	82	0.00
34 I	Perylene-d12	0.400	0.400	0.0	86	0.00
35	Benzo(b)fluoranthene	0.400	0.376	6.0	84	0.00
36	Benzo(k)fluoranthene	0.400	0.386	3.5	87	0.00
37 C	Benzo(a)pyrene	0.400	0.376	6.0	83	0.00
38	Dibenzo(a,h)anthracene	0.400	0.365	8.8	81	0.00
39	Benzo(a,h,i)perylene	0.400	0.370	7.5	83	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121218\
Data File : BE098362.D
Acq On : 12 Dec 2018 15:42
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 13 05:51:55 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			