

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121218\
 Data File : BE098362.D
 Acq On : 12 Dec 2018 15:42
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 13 05:51:55 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
2	1,4-Dioxane	0.692	0.718	-3.8	85	0.00
3	n-Nitrosodimethylamine	0.623	0.419	32.7#	58	0.00
4 S	2-Fluorophenol	0.936	0.897	4.2	80	0.01
5 S	Phenol-d6	1.326	1.204	9.2	73	0.00
6	bis(2-Chloroethyl)ether	1.260	1.052	16.5	70	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	78	0.00
8 S	Nitrobenzene-d5	0.364	0.348	4.4	82	0.01
9	Naphthalene	4.025	3.858	4.1	77	0.00
10	Hexachlorobutadiene	0.256	0.251	2.0	79	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.616	5.2	75	0.00
12	2-Methylnaphthalene	0.654	0.593	9.3	73	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	74	0.00
14 S	2,4,6-Tribromophenol	0.147	0.125	15.0	66	0.00
15 S	2-Fluorobiphenyl	1.687	1.532	9.2	74	0.00
16	Acenaphthylene	1.874	1.798	4.1	73	0.00
17	Acenaphthene	1.126	1.051	6.7	70	0.00
18	Fluorene	1.506	1.437	4.6	71	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	71	0.00
20	4-Bromophenyl-phenylether	0.215	0.194	9.8	67	0.00
21	Hexachlorobenzene	0.231	0.222	3.9	70	0.00
22	Pentachlorophenol	0.046	0.033	28.3#	65	0.01
23	Phenanthrene	0.972	0.900	7.4	67	0.00
24	Anthracene	0.866	0.760	12.2	63	0.00
25 SURR	Fluoranthene-d10	5.131	4.907	4.4	65	0.00
26	Fluoranthene	1.286	1.244	3.3	66	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
28	Pvrene	1.255	1.061	15.5	67	0.00
29 S	Terphenyl-d14	0.972	0.767	21.1	65	0.00
30	Benzo(a)anthracene	1.139	1.004	11.9	73	0.00
31	Chrysene	1.193	1.140	4.4	79	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	1.968	14.9	70	0.00
33	Indeno(1,2,3-cd)pyrene	1.187	1.135	4.4	82	0.00
34 I	Perylene-d12	1.000	1.000	0.0	86	0.00
35	Benzo(b)fluoranthene	1.094	1.029	5.9	84	0.00
36	Benzo(k)fluoranthene	1.274	1.230	3.5	87	0.00
37 C	Benzo(a)pyrene	1.129	1.061	6.0	83	0.00
38	Dibenzo(a,h)anthracene	1.063	0.970	8.7	81	0.00
39	Benzo(a,h,i)perylene	1.071	0.992	7.4	83	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121218\
Data File : BE098362.D
Acq On : 12 Dec 2018 15:42
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 13 05:51:55 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			