

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121218\
 Data File : BE098379.D
 Acq On : 13 Dec 2018 2:11
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Dec 13 07:25:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2	1,4-Dioxane	0.692	0.642	7.2	73	0.00
3	n-Nitrosodimethylamine	0.623	0.395	36.6#	53	0.00
4 S	2-Fluorophenol	0.936	0.858	8.3	73	0.01
5 S	Phenol-d6	1.326	1.208	8.9	70	0.00
6	bis(2-Chloroethyl)ether	1.260	1.016	19.4	64	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	76	-0.01
8 S	Nitrobenzene-d5	0.364	0.348	4.4	79	0.01
9	Naphthalene	4.025	3.948	1.9	76	0.00
10	Hexachlorobutadiene	0.256	0.268	-4.7	82	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.629	3.2	74	0.00
12	2-Methylnaphthalene	0.654	0.609	6.9	72	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	76	0.00
14 S	2,4,6-Tribromophenol	0.147	0.145	1.4	79	0.00
15 S	2-Fluorobiphenyl	1.687	1.523	9.7	76	0.00
16	Acenaphthylene	1.874	1.822	2.8	77	0.00
17	Acenaphthene	1.126	1.081	4.0	74	0.00
18	Fluorene	1.506	1.438	4.5	73	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	74	-0.01
20	4-Bromophenyl-phenylether	0.215	0.190	11.6	68	0.00
21	Hexachlorobenzene	0.231	0.224	3.0	73	0.00
22	Pentachlorophenol	0.046	0.027	41.3#	56	0.00
23	Phenanthrene	0.972	0.907	6.7	70	0.00
24	Anthracene	0.866	0.743	14.2	64	0.00
25 SURR	Fluoranthene-d10	5.131	4.997	2.6	70	0.00
26	Fluoranthene	1.286	1.279	0.5	71	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	84	0.00
28	Pvrene	1.255	1.089	13.2	71	0.00
29 S	Terphenyl-d14	0.972	0.787	19.0	69	0.00
30	Benzo(a)anthracene	1.139	1.047	8.1	79	-0.01
31	Chrysene	1.193	1.124	5.8	80	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	1.874	19.0	69	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.158	2.4	87	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	90	0.00
35	Benzo(b)fluoranthene	1.094	1.072	2.0	91	0.00
36	Benzo(k)fluoranthene	1.274	1.268	0.5	93	-0.01
37 C	Benzo(a)pyrene	1.129	1.104	2.2	90	0.00
38	Dibenzo(a,h)anthracene	1.063	0.994	6.5	86	0.00
39	Benzo(a,h,i)perylene	1.071	1.011	5.6	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121218\
Data File : BE098379.D
Acq On : 13 Dec 2018 2:11
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 13 07:25:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		