

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121218\
 Data File : BE098381.D
 Acq On : 13 Dec 2018 3:25
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 13 07:26:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	81	0.00
2	1,4-Dioxane	0.400	0.368	8.0	76	0.00
3	n-Nitrosodimethylamine	0.400	0.248	38.0#	54	0.01
4 S	2-Fluorophenol	0.400	0.372	7.0	78	0.01
5 S	Phenol-d6	0.400	0.375	6.3	76	0.00
6	bis(2-Chloroethyl)ether	0.400	0.329	17.8	69	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	81	-0.01
8 S	Nitrobenzene-d5	0.400	0.378	5.5	84	0.01
9	Naphthalene	0.400	0.389	2.8	81	0.00
10	Hexachlorobutadiene	0.400	0.421	-5.2	88	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.378	5.5	78	0.00
12	2-Methylnaphthalene	0.400	0.369	7.8	76	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	82	0.00
14 S	2,4,6-Tribromophenol	0.400	0.382	4.5	83	0.00
15 S	2-Fluorobiphenyl	0.400	0.360	10.0	81	0.00
16	Acenaphthylene	0.400	0.369	7.8	78	0.00
17	Acenaphthene	0.400	0.372	7.0	77	0.00
18	Fluorene	0.400	0.370	7.5	76	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	78	-0.01
20	4-Bromophenyl-phenylether	0.400	0.351	12.3	72	0.00
21	Hexachlorobenzene	0.400	0.378	5.5	75	0.00
22	Pentachlorophenol	0.400	0.300	25.0#	58	0.00
23	Phenanthrene	0.400	0.371	7.3	74	0.00
24	Anthracene	0.400	0.340	15.0	67	0.00
25 SURR	Fluoranthene-d10	0.400	0.382	4.5	72	0.00
26	Fluoranthene	0.400	0.388	3.0	73	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	88	0.00
28	Pvrene	0.400	0.339	15.3	72	0.00
29 S	Terphenyl-d14	0.400	0.324	19.0	72	0.00
30	Benzo(a)anthracene	0.400	0.359	10.3	80	-0.01
31	Chrysene	0.400	0.376	6.0	83	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.330	17.5	73	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.390	2.5	90	0.00
34 I	Perylene-d12	0.400	0.400	0.0	93	-0.01
35	Benzo(b)fluoranthene	0.400	0.363	9.3	87	0.00
36	Benzo(k)fluoranthene	0.400	0.386	3.5	94	0.00
37 C	Benzo(a)pyrene	0.400	0.366	8.5	87	0.00
38	Dibenzo(a,h)anthracene	0.400	0.368	8.0	88	-0.01
39	Benzo(a,h,i)perylene	0.400	0.376	6.0	90	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121218\
Data File : BE098381.D
Acq On : 13 Dec 2018 3:25
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 13 07:26:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				