

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121218\
 Data File : BE098381.D
 Acq On : 13 Dec 2018 3:25
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 13 07:26:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
2	1,4-Dioxane	0.692	0.636	8.1	76	0.00
3	n-Nitrosodimethylamine	0.623	0.386	38.0#	54	0.01
4 S	2-Fluorophenol	0.936	0.871	6.9	78	0.01
5 S	Phenol-d6	1.326	1.243	6.3	76	0.00
6	bis(2-Chloroethyl)ether	1.260	1.036	17.8	69	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	81	-0.01
8 S	Nitrobenzene-d5	0.364	0.344	5.5	84	0.01
9	Naphthalene	4.025	3.912	2.8	81	0.00
10	Hexachlorobutadiene	0.256	0.270	-5.5	88	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.615	5.4	78	0.00
12	2-Methylnaphthalene	0.654	0.603	7.8	76	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	82	0.00
14 S	2,4,6-Tribromophenol	0.147	0.140	4.8	83	0.00
15 S	2-Fluorobiphenyl	1.687	1.520	9.9	81	0.00
16	Acenaphthylene	1.874	1.729	7.7	78	0.00
17	Acenaphthene	1.126	1.046	7.1	77	0.00
18	Fluorene	1.506	1.392	7.6	76	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	78	-0.01
20	4-Bromophenyl-phenylether	0.215	0.189	12.1	72	0.00
21	Hexachlorobenzene	0.231	0.218	5.6	75	0.00
22	Pentachlorophenol	0.046	0.027	41.3#	58	0.00
23	Phenanthrene	0.972	0.902	7.2	74	0.00
24	Anthracene	0.866	0.736	15.0	67	0.00
25 SURR	Fluoranthene-d10	5.131	4.897	4.6	72	0.00
26	Fluoranthene	1.286	1.248	3.0	73	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
28	Pvrene	1.255	1.062	15.4	72	0.00
29 S	Terphenyl-d14	0.972	0.788	18.9	72	0.00
30	Benzo(a)anthracene	1.139	1.023	10.2	80	-0.01
31	Chrysene	1.193	1.123	5.9	83	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	1.909	17.5	73	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.159	2.4	90	0.00
34 I	Perylene-d12	1.000	1.000	0.0	93	-0.01
35	Benzo(b)fluoranthene	1.094	0.994	9.1	87	0.00
36	Benzo(k)fluoranthene	1.274	1.231	3.4	94	0.00
37 C	Benzo(a)pyrene	1.129	1.034	8.4	87	0.00
38	Dibenzo(a,h)anthracene	1.063	0.979	7.9	88	-0.01
39	Benzo(a,h,i)perylene	1.071	1.007	6.0	90	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121218\
Data File : BE098381.D
Acq On : 13 Dec 2018 3:25
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 13 07:26:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			