

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121218\
 Data File : BE098395.D
 Acq On : 13 Dec 2018 14:01
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 13 14:48:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	83	0.00
2	1,4-Dioxane	0.400	0.380	5.0	80	0.00
3	n-Nitrosodimethylamine	0.400	0.255	36.3#	57	0.01
4 S	2-Fluorophenol	0.400	0.352	12.0	75	0.01
5 S	Phenol-d6	0.400	0.370	7.5	76	0.00
6	bis(2-Chloroethyl)ether	0.400	0.305	23.8	65	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	82	-0.01
8 S	Nitrobenzene-d5	0.400	0.352	12.0	79	0.01
9	Naphthalene	0.400	0.389	2.8	82	0.00
10	Hexachlorobutadiene	0.400	0.415	-3.7	88	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.375	6.3	78	0.00
12	2-Methylnaphthalene	0.400	0.364	9.0	76	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	81	0.00
14 S	2,4,6-Tribromophenol	0.400	0.337	15.8	72	0.00
15 S	2-Fluorobiphenyl	0.400	0.361	9.8	80	0.00
16	Acenaphthylene	0.400	0.377	5.8	79	0.00
17	Acenaphthene	0.400	0.372	7.0	76	0.00
18	Fluorene	0.400	0.367	8.3	75	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	75	-0.01
20	4-Bromophenyl-phenylether	0.400	0.353	11.8	69	0.00
21	Hexachlorobenzene	0.400	0.372	7.0	71	0.00
22	Pentachlorophenol	0.400	0.279	30.3#	51	0.00
23	Phenanthrene	0.400	0.375	6.3	71	0.00
24	Anthracene	0.400	0.331	17.3	63	0.00
25 SURR	Fluoranthene-d10	0.400	0.368	8.0	67	0.00
26	Fluoranthene	0.400	0.379	5.3	68	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	83	0.00
28	Pvrene	0.400	0.344	14.0	70	0.00
29 S	Terphenyl-d14	0.400	0.316	21.0	66	0.00
30	Benzo(a)anthracene	0.400	0.356	11.0	76	-0.01
31	Chrysene	0.400	0.380	5.0	80	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.286	28.5#	60	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.392	2.0	86	0.00
34 I	Perylene-d12	0.400	0.400	0.0	90	-0.01
35	Benzo(b)fluoranthene	0.400	0.370	7.5	86	0.00
36	Benzo(k)fluoranthene	0.400	0.408	-2.0	95	0.00
37 C	Benzo(a)pyrene	0.400	0.384	4.0	88	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.369	7.8	85	0.00
39	Benzo(a,h,i)perylene	0.400	0.377	5.8	87	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121218\
Data File : BE098395.D
Acq On : 13 Dec 2018 14:01
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 13 14:48:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				