

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121318\
 Data File : BE098403.D
 Acq On : 13 Dec 2018 18:58
 Operator : SJ/JU
 Sample : J6323-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RE-122D1-20181206

Quant Time: Dec 13 21:10:30 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.87	152	938	0.40	ng	0.00
7) Naphthalene-d8	10.67	136	3917	0.40	ng	0.00
13) Acenaphthene-d10	14.53	164	2459	0.40	ng	0.00
19) Phenanthrene-d10	17.27	188	7034	0.40	ng	-0.01
27) Chrysene-d12	21.46	240	8459	0.40	ng	0.00
34) Perylene-d12	24.00	264	7934	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.46	112	439	0.20	ng	0.01
5) Phenol-d6	7.04	99	375	0.12	ng	0.00
8) Nitrobenzene-d5	9.03	82	1405	0.39	ng	0.01
11) 2-Methylnaphthalene-d10	12.27	152	2709	0.43	ng	0.00
14) 2,4,6-Tribromophenol	16.02	330	552	0.61	ng	0.00
15) 2-Fluorobiphenyl	13.15	172	4847	0.47	ng	0.00
25) Fluoranthene-d10	19.30	212	41683	0.46	ng	0.00
29) Terphenyl-d14	19.90	244	7894	0.38	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.36	88	5091	3.138	ng	# 93
26) Fluoranthene	19.33	202	494	0.022	ng	# 95
32) Bis(2-ethylhexyl)phthalate	21.35	149	3339	0.068	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121318\
Data File : BE098403.D
Acq On : 13 Dec 2018 18:58
Operator : SJ/JU
Sample : J6323-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RE-122D1-20181206

Quant Time: Dec 13 21:10:30 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

