


Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121318\
 Data File : BE098396.D
 Acq On : 13 Dec 2018 14:40
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121018.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Dec 10 14:59:55 2018



AutoFind: Scans 2573, 2574, 2575; Background Corrected with Scan 2561

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	60.7	12771	PASS
68	69	0.00	2	0.7	76	PASS
69	198	0.00	100	52.6	11063	PASS
70	69	0.00	2	0.7	76	PASS
127	198	10	80	50.7	10674	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	21036	PASS
199	198	5	9	6.5	1368	PASS
275	198	10	60	28.8	6052	PASS
365	198	1	100	4.4	919	PASS
441	198	0.01	100	13.4	2809	PASS
442	198	50	100	75.0	15768	PASS
443	442	15	24	18.2	2865	PASS