

Data Path : Z:\HPCHEM1\BNA E\Data\BE121417\
 Data File : BE094982.D
 Acq On : 14 Dec 2017 18:25
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE121417

Quant Time: Dec 14 19:07:14 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 14 18:42:53 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.394	0.376	4.6	83	0.00
3	n-Nitrosodimethylamine	0.466	0.456	2.1	90	0.00
4 S	2-Fluorophenol	0.621	0.599	3.5	92	0.00
5 S	Phenol-d6	0.923	0.859	6.9	88	0.00
6	bis(2-Chloroethyl)ether	0.901	0.866	3.9	89	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
8 S	Nitrobenzene-d5	0.328	0.328	0.0	94	0.00
9	Naphthalene	4.697	4.744	-1.0	91	0.00
10	Hexachlorobutadiene	0.214	0.213	0.5	89	0.00
11 SURR	2-Methylnaphthalene-d10	0.655	0.665	-1.5	92	0.00
12	2-Methylnaphthalene	1.172	1.164	0.7	90	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.093	0.086	7.5	91	0.00
15 S	2-Fluorobiphenyl	1.772	1.753	1.1	89	0.00
16	Acenaphthylene	2.138	2.060	3.6	92	0.00
17	Acenaphthene	1.418	1.398	1.4	91	0.00
18	Fluorene	1.762	1.713	2.8	91	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
20	4-Bromophenyl-phenylether	0.234	0.233	0.4	92	0.00
21	Hexachlorobenzene	0.239	0.243	-1.7	93	0.00
22	Pentachlorophenol	0.046	0.040	13.0	82	0.00
23	Phenanthrene	1.246	1.244	0.2	92	0.00
24	Anthracene	1.279	1.245	2.7	93	0.00
25 SURR	Fluoranthene-d10	5.189	5.023	3.2	91	0.00
26	Fluoranthene	1.538	1.503	2.3	91	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
28	Pvrene	1.355	1.329	1.9	86	0.00
29 S	Terphenyl-d14	0.753	0.761	-1.1	91	0.00
30	Benzo(a)anthracene	1.169	1.122	4.0	89	0.00
31	Chrysene	1.316	1.327	-0.8	92	0.00
32	Bis(2-ethylhexyl)phthalate	1.489	1.232	17.3	84	0.00
33	Indeno(1,2,3-cd)pyrene	1.071	1.004	6.3	84	0.00
34 I	Perylene-d12	1.000	1.000	0.0	85	0.00
35	Benzo(b)fluoranthene	1.469	1.419	3.4	82	0.00
36	Benzo(k)fluoranthene	1.488	1.470	1.2	89	0.00
37 C	Benzo(a)pyrene	1.303	1.223	6.1	81	0.00
38	Dibenzo(a,h)anthracene	1.209	1.149	5.0	82	0.00
39	Benzo(a,h,i)perylene	1.203	1.192	0.9	86	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE121417\
Data File : BE094982.D
Acq On : 14 Dec 2017 18:25
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE121417

Quant Time: Dec 14 19:07:14 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 14 18:42:53 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			