

Data Path : Z:\HPCHEM1\BNA E\Data\BE121417\
 Data File : BE094993.D
 Acq On : 15 Dec 2017 00:56
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 15 08:14:23 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 14 18:42:53 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.394	0.404	-2.5	85	0.00
3	n-Nitrosodimethylamine	0.466	0.521	-11.8	98	0.00
4 S	2-Fluorophenol	0.621	0.702	-13.0	104	0.00
5 S	Phenol-d6	0.923	1.075	-16.5	106	0.00
6	bis(2-Chloroethyl)ether	0.901	0.935	-3.8	93	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.328	0.381	-16.2	116	0.00
9	Naphthalene	4.697	4.664	0.7	95	0.00
10	Hexachlorobutadiene	0.214	0.214	0.0	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.655	0.663	-1.2	97	0.00
12	2-Methylnaphthalene	1.172	1.174	-0.2	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.093	0.146	-57.0#	165#	0.00
15 S	2-Fluorobiphenyl	1.772	1.706	3.7	93	0.00
16	Acenaphthylene	2.138	2.146	-0.4	103	0.00
17	Acenaphthene	1.418	1.427	-0.6	100	0.00
18	Fluorene	1.762	1.805	-2.4	103	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.234	0.235	-0.4	103	0.00
21	Hexachlorobenzene	0.239	0.232	2.9	98	0.00
22	Pentachlorophenol	0.046	0.085	-84.8#	192#	0.00
23	Phenanthrene	1.246	1.221	2.0	100	-0.02
24	Anthracene	1.279	1.291	-0.9	106	-0.02
25 SURR	Fluoranthene-d10	5.189	5.427	-4.6	109	0.00
26	Fluoranthene	1.538	1.580	-2.7	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
28	Pvrene	1.355	1.370	-1.1	106	0.00
29 S	Terphenyl-d14	0.753	0.768	-2.0	109	0.00
30	Benzo(a)anthracene	1.169	1.248	-6.8	117	0.00
31	Chrysene	1.316	1.249	5.1	102	-0.02
32	Bis(2-ethylhexyl)phthalate	1.489	2.570	-72.6#	208#	-0.02
33	Indeno(1,2,3-cd)pyrene	1.071	1.133	-5.8	112	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.469	1.359	7.5	104	0.00
36	Benzo(k)fluoranthene	1.488	1.389	6.7	110	0.00
37 C	Benzo(a)pyrene	1.303	1.262	3.1	109	0.00
38	Dibenzo(a,h)anthracene	1.209	1.157	4.3	109	-0.01
39	Benzo(a,h,i)perylene	1.203	1.138	5.4	109	-0.01

Data Path : Z:\HPCHEM1\BNA E\Data\BE121417\
Data File : BE094993.D
Acq On : 15 Dec 2017 00:56
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 15 08:14:23 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 14 18:42:53 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			