

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121618\
 Data File : BE098504.D
 Acq On : 16 Dec 2018 13:14
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Dec 17 03:16:01 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	47#	0.00
2	1,4-Dioxane	0.692	0.783	-13.2	54	0.00
3	n-Nitrosodimethylamine	0.623	0.705	-13.2	58	0.00
4 S	2-Fluorophenol	0.936	1.195	-27.7#	62	0.00
5 S	Phenol-d6	1.326	1.528	-15.2	54	0.00
6	bis(2-Chloroethyl)ether	1.260	1.224	2.9	47#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	45#	-0.01
8 S	Nitrobenzene-d5	0.364	0.342	6.0	46#	0.00
9	Naphthalene	4.025	4.055	-0.7	46#	0.00
10	Hexachlorobutadiene	0.256	0.288	-12.5	52	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.672	-3.4	47#	-0.01
12	2-Methylnaphthalene	0.654	0.705	-7.8	50#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	44#	-0.01
14 S	2,4,6-Tribromophenol	0.147	0.220	-49.7#	70	-0.01
15 S	2-Fluorobiphenyl	1.687	1.743	-3.3	50	0.00
16	Acenaphthylene	1.874	1.977	-5.5	48#	0.00
17	Acenaphthene	1.126	1.188	-5.5	47#	0.00
18	Fluorene	1.506	1.636	-8.6	48#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	46#	-0.01
20	4-Bromophenyl-phenylether	0.215	0.223	-3.7	49#	-0.01
21	Hexachlorobenzene	0.231	0.248	-7.4	50	-0.01
22	Pentachlorophenol	0.046	0.114	-147.8#	144	-0.01
23	Phenanthrene	0.972	1.017	-4.6	49#	0.00
24	Anthracene	0.866	0.886	-2.3	47#	0.00
25 SURR	Fluoranthene-d10	5.131	6.167	-20.2	53	0.00
26	Fluoranthene	1.286	1.534	-19.3	52	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	60	0.00
28	Pvrene	1.255	1.216	3.1	57	0.00
29 S	Terphenyl-d14	0.972	0.882	9.3	55	0.00
30	Benzo(a)anthracene	1.139	1.242	-9.0	67	-0.01
31	Chrysene	1.193	1.216	-1.9	62	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	3.217	-39.1#	84	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.224	-3.1	65	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	63	-0.01
35	Benzo(b)fluoranthene	1.094	1.191	-8.9	70	-0.01
36	Benzo(k)fluoranthene	1.274	1.294	-1.6	66	-0.01
37 C	Benzo(a)pyrene	1.129	1.198	-6.1	68	-0.01
38	Dibenzo(a,h)anthracene	1.063	1.023	3.8	62	-0.01
39	Benzo(a,h,i)perylene	1.071	1.117	-4.3	67	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121618\
Data File : BE098504.D
Acq On : 16 Dec 2018 13:14
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 17 03:16:01 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			