

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121618\
 Data File : BE098507.D
 Acq On : 16 Dec 2018 15:09
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 03:25:49 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	43#	0.00
2	1,4-Dioxane	0.692	0.749	-8.2	48#	0.00
3	n-Nitrosodimethylamine	0.623	0.708	-13.6	53	0.01
4 S	2-Fluorophenol	0.936	1.148	-22.6	55	0.00
5 S	Phenol-d6	1.326	1.490	-12.4	48#	0.00
6	bis(2-Chloroethyl)ether	1.260	1.116	11.4	40#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	42#	-0.01
8 S	Nitrobenzene-d5	0.364	0.325	10.7	41#	0.01
9	Naphthalene	4.025	4.192	-4.1	45#	0.00
10	Hexachlorobutadiene	0.256	0.312	-21.9	53	-0.01
11 SURR	2-Methylnaphthalene-d10	0.650	0.681	-4.8	45#	-0.01
12	2-Methylnaphthalene	0.654	0.659	-0.8	43#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	43#	-0.01
14 S	2,4,6-Tribromophenol	0.147	0.193	-31.3#	60	0.00
15 S	2-Fluorobiphenyl	1.687	1.651	2.1	47#	0.00
16	Acenaphthylene	1.874	1.939	-3.5	47#	0.00
17	Acenaphthene	1.126	1.140	-1.2	44#	0.00
18	Fluorene	1.506	1.612	-7.0	47#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	45#	-0.01
20	4-Bromophenyl-phenylether	0.215	0.199	7.4	44#	0.00
21	Hexachlorobenzene	0.231	0.250	-8.2	50	0.00
22	Pentachlorophenol	0.046	0.072	-56.5#	90	-0.01
23	Phenanthrene	0.972	0.995	-2.4	47#	0.00
24	Anthracene	0.866	0.792	8.5	42#	0.00
25 SURR	Fluoranthene-d10	5.131	5.830	-13.6	50#	0.00
26	Fluoranthene	1.286	1.476	-14.8	50#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	58	-0.01
28	Pvrene	1.255	1.202	4.2	54	0.00
29 S	Terphenyl-d14	0.972	0.831	14.5	50#	0.00
30	Benzo(a)anthracene	1.139	1.196	-5.0	62	-0.01
31	Chrysene	1.193	1.188	0.4	58	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	1.844	20.3	46#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.221	-2.9	62	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	59	-0.01
35	Benzo(b)fluoranthene	1.094	1.155	-5.6	65	-0.01
36	Benzo(k)fluoranthene	1.274	1.394	-9.4	68	-0.01
37 C	Benzo(a)pyrene	1.129	1.200	-6.3	64	-0.01
38	Dibenzo(a,h)anthracene	1.063	1.050	1.2	60	-0.01
39	Benzo(a,h,i)perylene	1.071	1.159	-8.2	66	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121618\
Data File : BE098507.D
Acq On : 16 Dec 2018 15:09
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 17 03:25:49 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			