

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121619\
 Data File : BE100915.D
 Acq On : 16 Dec 2019 9:58
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 16 10:51:50 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 13 10:42:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.729	0.798	-9.5	103	0.00
3	n-Nitrosodimethylamine	0.594	0.615	-3.5	112	0.00
4 S	2-Fluorophenol	0.949	1.124	-18.4	119	0.00
5 S	Phenol-d6	1.433	1.676	-17.0	119	0.00
6	bis(2-Chloroethyl)ether	1.385	1.483	-7.1	107	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.190	0.287	-51.1#	170#	0.00
9	Naphthalene	4.250	4.585	-7.9	110	0.00
10	Hexachlorobutadiene	0.166	0.173	-4.2	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.628	-8.3	113	0.00
12	2-Methylnaphthalene	0.664	0.723	-8.9	113	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.044	0.072	-63.6#	189#	-0.01
15 S	2-Fluorobiphenyl	1.587	1.550	2.3	104	0.00
16	Acenaphthylene	1.655	1.789	-8.1	127	0.00
17	Acenaphthene	1.140	1.195	-4.8	117	0.00
18	Fluorene	1.471	1.546	-5.1	120	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
20	4-Bromophenyl-phenylether	0.196	0.213	-8.7	112	0.00
21	Hexachlorobenzene	0.213	0.232	-8.9	108	0.00
22	Pentachlorophenol	0.015	0.024	-60.0#	779#	-0.01
23	Phenanthrene	1.212	1.284	-5.9	105	0.00
24	Anthracene	1.013	1.022	-0.9	110	0.00
25 SURR	Fluoranthene-d10	4.728	4.214	10.9	92	0.00
26	Fluoranthene	1.446	1.325	8.4	95	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	74	-0.01
28	Pvrene	1.300	1.557	-19.8	95	0.00
29 S	Terphenyl-d14	0.958	1.041	-8.7	84	0.00
30	Benzo(a)anthracene	1.179	1.309	-11.0	93	0.00
31	Chrysene	1.385	1.576	-13.8	87	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	1.421	-32.9#	130	0.00
33	Indeno(1,2,3-cd)pyrene	1.207	1.317	-9.1	92	0.00
34 I	Perylene-d12	1.000	1.000	0.0	78	0.00
35	Benzo(b)fluoranthene	1.095	1.185	-8.2	89	0.00
36	Benzo(k)fluoranthene	1.228	1.341	-9.2	89	0.00
37 C	Benzo(a)pyrene	0.948	1.031	-8.8	93	0.00
38	Dibenzo(a,h)anthracene	0.943	0.974	-3.3	90	0.00
39	Benzo(a,h,i)perylene	1.009	1.069	-5.9	88	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121619\
Data File : BE100915.D
Acq On : 16 Dec 2019 9:58
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 16 10:51:50 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 13 10:42:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			