

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121619\
 Data File : BE100928.D
 Acq On : 16 Dec 2019 20:05
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 06:19:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 13 10:42:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	84	0.00
2	1,4-Dioxane	0.400	0.444	-11.0	93	0.00
3	n-Nitrosodimethylamine	0.400	0.406	-1.5	98	0.00
4 S	2-Fluorophenol	0.400	0.484	-21.0	108	-0.01
5 S	Phenol-d6	0.400	0.469	-17.2	106	0.00
6	bis(2-Chloroethyl)ether	0.400	0.438	-9.5	97	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	87	0.00
8 S	Nitrobenzene-d5	0.400	0.605	-51.2#	151	0.00
9	Naphthalene	0.400	0.432	-8.0	98	0.00
10	Hexachlorobutadiene	0.400	0.425	-6.2	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.431	-7.7	99	-0.01
12	2-Methylnaphthalene	0.400	0.441	-10.2	102	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	93	0.00
14 S	2,4,6-Tribromophenol	0.400	0.727	-81.7#	194	-0.01
15 S	2-Fluorobiphenyl	0.400	0.387	3.3	94	0.00
16	Acenaphthylene	0.400	0.435	-8.7	117	0.00
17	Acenaphthene	0.400	0.427	-6.7	109	-0.01
18	Fluorene	0.400	0.431	-7.7	113	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	91	0.00
20	4-Bromophenyl-phenylether	0.400	0.410	-2.5	103	0.00
21	Hexachlorobenzene	0.400	0.415	-3.7	100	0.00
22	Pentachlorophenol	0.400	0.814	-103.5#	957	-0.01
23	Phenanthrene	0.400	0.422	-5.5	102	0.00
24	Anthracene	0.400	0.410	-2.5	108	0.00
25 SURR	Fluoranthene-d10	0.400	0.363	9.3	91	0.00
26	Fluoranthene	0.400	0.371	7.3	93	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	74	-0.01
28	Pvrene	0.400	0.477	-19.2	95	0.00
29 S	Terphenyl-d14	0.400	0.424	-6.0	82	0.00
30	Benzo(a)anthracene	0.400	0.459	-14.8	96	0.00
31	Chrysene	0.400	0.450	-12.5	86	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.749	-87.3#	147	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.464	-16.0	98	0.00
34 I	Perylene-d12	0.400	0.400	0.0	85	0.00
35	Benzo(b)fluoranthene	0.400	0.415	-3.7	92	0.00
36	Benzo(k)fluoranthene	0.400	0.413	-3.2	92	0.00
37 C	Benzo(a)pyrene	0.400	0.416	-4.0	97	0.00
38	Dibenzo(a,h)anthracene	0.400	0.406	-1.5	97	0.00
39	Benzo(a,h,i)perylene	0.400	0.418	-4.5	94	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121619\
Data File : BE100928.D
Acq On : 16 Dec 2019 20:05
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 17 06:19:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 13 10:42:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				