

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121619\
 Data File : BE100930.D
 Acq On : 16 Dec 2019 22:34
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 14:27:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 13 10:42:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.729	0.851	-16.7	101	0.01
3	n-Nitrosodimethylamine	0.594	0.571	3.9	95	0.01
4 S	2-Fluorophenol	0.949	1.096	-15.5	106	-0.01
5 S	Phenol-d6	1.433	1.623	-13.3	105	0.00
6	bis(2-Chloroethyl)ether	1.385	1.505	-8.7	99	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
8 S	Nitrobenzene-d5	0.190	0.266	-40.0#	143	0.00
9	Naphthalene	4.250	4.656	-9.6	102	0.00
10	Hexachlorobutadiene	0.166	0.177	-6.6	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.634	-9.3	104	-0.01
12	2-Methylnaphthalene	0.664	0.730	-9.9	104	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.044	0.075	-70.5#	184#	-0.01
15 S	2-Fluorobiphenyl	1.587	1.575	0.8	99	0.00
16	Acenaphthylene	1.655	1.764	-6.6	118	0.00
17	Acenaphthene	1.140	1.245	-9.2	114	0.00
18	Fluorene	1.471	1.614	-9.7	118	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	4-Bromophenyl-phenylether	0.196	0.204	-4.1	109	0.00
21	Hexachlorobenzene	0.213	0.232	-8.9	110	0.00
22	Pentachlorophenol	0.015	0.025	-66.7#	832#	-0.01
23	Phenanthrene	1.212	1.275	-5.2	106	0.00
24	Anthracene	1.013	0.991	2.2	108	0.00
25 SURR	Fluoranthene-d10	4.728	4.038	14.6	90	0.00
26	Fluoranthene	1.446	1.302	10.0	95	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	72	-0.01
28	Pvrene	1.300	1.544	-18.8	92	0.00
29 S	Terphenyl-d14	0.958	1.034	-7.9	82	0.00
30	Benzo(a)anthracene	1.179	1.281	-8.7	88	0.00
31	Chrysene	1.385	1.566	-13.1	84	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	1.255	-17.4	112	0.00
33	Indeno(1,2,3-cd)pyrene	1.207	1.301	-7.8	88	0.00
34 I	Perylene-d12	1.000	1.000	0.0	81	0.00
35	Benzo(b)fluoranthene	1.095	1.083	1.1	84	0.00
36	Benzo(k)fluoranthene	1.228	1.298	-5.7	90	0.00
37 C	Benzo(a)pyrene	0.948	0.888	6.3	83	0.00
38	Dibenzo(a,h)anthracene	0.943	0.914	3.1	88	0.00
39	Benzo(a,h,i)perylene	1.009	0.995	1.4	85	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121619\
Data File : BE100930.D
Acq On : 16 Dec 2019 22:34
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 17 14:27:29 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 13 10:42:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			