

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121619\
 Data File : BE100944.D
 Acq On : 17 Dec 2019 8:51
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 12:22:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 13 10:42:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	0.00
2	1,4-Dioxane	0.400	0.408	-2.0	97	0.00
3	n-Nitrosodimethylamine	0.400	0.479	-19.7	131	0.00
4 S	2-Fluorophenol	0.400	0.628	-57.0#	160	0.00
5 S	Phenol-d6	0.400	0.619	-54.7#	160	0.00
6	bis(2-Chloroethyl)ether	0.400	0.478	-19.5	120	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	105	0.00
8 S	Nitrobenzene-d5	0.400	0.727	-81.7#	219	0.00
9	Naphthalene	0.400	0.442	-10.5	121	0.00
10	Hexachlorobutadiene	0.400	0.410	-2.5	111	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.453	-13.2	127	0.00
12	2-Methylnaphthalene	0.400	0.466	-16.5	130	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	125	0.00
14 S	2,4,6-Tribromophenol	0.400	1.068	-167.0#	383	0.00
15 S	2-Fluorobiphenyl	0.400	0.366	8.5	120	0.00
16	Acenaphthylene	0.400	0.475	-18.7	172	0.00
17	Acenaphthene	0.400	0.431	-7.7	148	0.00
18	Fluorene	0.400	0.448	-12.0	158	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	130	0.00
20	4-Bromophenyl-phenylether	0.400	0.420	-5.0	151	0.00
21	Hexachlorobenzene	0.400	0.376	6.0	130	0.00
22	Pentachlorophenol	0.400	0.865	-116.2#	1452	0.00
23	Phenanthrene	0.400	0.405	-1.3	140	0.00
24	Anthracene	0.400	0.448	-12.0	170	0.00
25 SURR	Fluoranthene-d10	0.400	0.405	-1.3	145	0.00
26	Fluoranthene	0.400	0.395	1.3	142	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	115	-0.01
28	Pvrene	0.400	0.488	-22.0	150	0.00
29 S	Terphenyl-d14	0.400	0.429	-7.2	129	0.00
30	Benzo(a)anthracene	0.400	0.516	-29.0#	167	0.00
31	Chrysene	0.400	0.443	-10.7	132	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	1.574	-293.5#	582	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.543	-35.8#	177	0.00
34 I	Perylene-d12	0.400	0.400	0.0	131	0.00
35	Benzo(b)fluoranthene	0.400	0.457	-14.2	158	0.00
36	Benzo(k)fluoranthene	0.400	0.434	-8.5	150	0.00
37 C	Benzo(a)pyrene	0.400	0.508	-27.0#	183	0.00
38	Dibenzo(a,h)anthracene	0.400	0.473	-18.2	175	0.00
39	Benzo(a,h,i)perylene	0.400	0.461	-15.3	160	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121619\
Data File : BE100944.D
Acq On : 17 Dec 2019 8:51
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 17 12:22:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 13 10:42:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 1				