

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121619\
 Data File : BE100944.D
 Acq On : 17 Dec 2019 8:51
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 12:22:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 13 10:42:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.729	0.743	-1.9	97	0.00
3	n-Nitrosodimethylamine	0.594	0.711	-19.7	131	0.00
4 S	2-Fluorophenol	0.949	1.491	-57.1#	160#	0.00
5 S	Phenol-d6	1.433	2.217	-54.7#	160#	0.00
6	bis(2-Chloroethyl)ether	1.385	1.655	-19.5	120	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.190	0.346	-82.1#	219#	0.00
9	Naphthalene	4.250	4.693	-10.4	121	0.00
10	Hexachlorobutadiene	0.166	0.170	-2.4	111	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.657	-13.3	127	0.00
12	2-Methylnaphthalene	0.664	0.773	-16.4	130	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.00
14 S	2,4,6-Tribromophenol	0.044	0.119	-170.5#	383#	0.00
15 S	2-Fluorobiphenyl	1.587	1.453	8.4	120	0.00
16	Acenaphthylene	1.655	1.965	-18.7	172#	0.00
17	Acenaphthene	1.140	1.230	-7.9	148	0.00
18	Fluorene	1.471	1.649	-12.1	158#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
20	4-Bromophenyl-phenylether	0.196	0.206	-5.1	151#	0.00
21	Hexachlorobenzene	0.213	0.200	6.1	130	0.00
22	Pentachlorophenol	0.015	0.032	-113.3#	1452#	0.00
23	Phenanthrene	1.212	1.226	-1.2	140	0.00
24	Anthracene	1.013	1.136	-12.1	170#	0.00
25 SURR	Fluoranthene-d10	4.728	4.782	-1.1	145	0.00
26	Fluoranthene	1.446	1.426	1.4	142	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	115	-0.01
28	Pvrene	1.300	1.586	-22.0	150#	0.00
29 S	Terphenyl-d14	0.958	1.028	-7.3	129	0.00
30	Benzo(a)anthracene	1.179	1.521	-29.0#	167#	0.00
31	Chrysene	1.385	1.534	-10.8	132	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	4.096	-283.2#	582#	0.00
33	Indeno(1,2,3-cd)pyrene	1.207	1.639	-35.8#	177#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	131	0.00
35	Benzo(b)fluoranthene	1.095	1.252	-14.3	158#	0.00
36	Benzo(k)fluoranthene	1.228	1.332	-8.5	150	0.00
37 C	Benzo(a)pyrene	0.948	1.204	-27.0#	183#	0.00
38	Dibenzo(a,h)anthracene	0.943	1.116	-18.3	175#	0.00
39	Benzo(a,h,i)perylene	1.009	1.162	-15.2	160#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121619\
Data File : BE100944.D
Acq On : 17 Dec 2019 8:51
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 17 12:22:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 13 10:42:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 1			