

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121918\
 Data File : BE098522.D
 Acq On : 19 Dec 2018 9:41
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 19 11:01:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	88	-0.01
2	1,4-Dioxane	0.400	0.348	13.0	78	-0.01
3	n-Nitrosodimethylamine	0.400	0.249	37.8#	59	0.00
4 S	2-Fluorophenol	0.400	0.400	0.0	91	0.00
5 S	Phenol-d6	0.400	0.483	-20.7	105	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.358	10.5	81	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	99	-0.01
8 S	Nitrobenzene-d5	0.400	0.348	13.0	94	0.00
9	Naphthalene	0.400	0.405	-1.3	103	-0.01
10	Hexachlorobutadiene	0.400	0.398	0.5	101	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.402	-0.5	101	-0.01
12	2-Methylnaphthalene	0.400	0.405	-1.3	102	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	96	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.373	6.8	94	-0.01
15 S	2-Fluorobiphenyl	0.400	0.391	2.3	103	-0.01
16	Acenaphthylene	0.400	0.363	9.3	90	-0.01
17	Acenaphthene	0.400	0.396	1.0	96	-0.01
18	Fluorene	0.400	0.380	5.0	92	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	80	-0.02
20	4-Bromophenyl-phenylether	0.400	0.401	-0.3	83	-0.01
21	Hexachlorobenzene	0.400	0.433	-8.2	88	-0.01
22	Pentachlorophenol	0.400	0.402	-0.5	82	-0.01
23	Phenanthrene	0.400	0.392	2.0	79	-0.01
24	Anthracene	0.400	0.360	10.0	73	-0.01
25 SURR	Fluoranthene-d10	0.400	0.387	3.3	74	-0.01
26	Fluoranthene	0.400	0.381	4.8	73	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	86	-0.01
28	Pvrene	0.400	0.364	9.0	77	-0.01
29 S	Terphenyl-d14	0.400	0.353	11.8	77	-0.01
30	Benzo(a)anthracene	0.400	0.377	5.8	83	-0.03
31	Chrysene	0.400	0.380	5.0	83	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.319	20.3	69	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.281	29.8#	64	-0.03
34 I	Perylene-d12	0.400	0.400	0.0	89	-0.03
35	Benzo(b)fluoranthene	0.400	0.386	3.5	89	-0.02
36	Benzo(k)fluoranthene	0.400	0.413	-3.2	96	-0.03
37 C	Benzo(a)pyrene	0.400	0.390	2.5	89	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.288	28.0#	66	-0.03
39	Benzo(a,h,i)perylene	0.400	0.263	34.3#	61	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121918\
Data File : BE098522.D
Acq On : 19 Dec 2018 9:41
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 19 11:01:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				