

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121918\
 Data File : BE098530.D
 Acq On : 19 Dec 2018 17:15
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 19 23:44:05 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.692	0.559	19.2	80	-0.01
3	n-Nitrosodimethylamine	0.623	0.540	13.3	91	0.00
4 S	2-Fluorophenol	0.936	1.006	-7.5	107	0.01
5 S	Phenol-d6	1.326	1.543	-16.4	112	0.00
6	bis(2-Chloroethyl)ether	1.260	1.110	11.9	88	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.364	0.327	10.2	109	0.01
9	Naphthalene	4.025	3.953	1.8	112	0.01
10	Hexachlorobutadiene	0.256	0.247	3.5	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.655	-0.8	113	0.00
12	2-Methylnaphthalene	0.654	0.645	1.4	112	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.147	0.149	-1.4	113	0.00
15 S	2-Fluorobiphenyl	1.687	1.662	1.5	114	0.00
16	Acenaphthylene	1.874	1.709	8.8	99	0.00
17	Acenaphthene	1.126	1.108	1.6	105	0.00
18	Fluorene	1.506	1.402	6.9	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	91	-0.01
20	4-Bromophenyl-phenylether	0.215	0.211	1.9	93	0.00
21	Hexachlorobenzene	0.231	0.240	-3.9	96	0.00
22	Pentachlorophenol	0.046	0.042	8.7	104	0.00
23	Phenanthrene	0.972	0.949	2.4	90	0.00
24	Anthracene	0.866	0.798	7.9	84	0.00
25 SURR	Fluoranthene-d10	5.131	4.782	6.8	81	0.00
26	Fluoranthene	1.286	1.189	7.5	80	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	-0.01
28	Pvrene	1.255	1.132	9.8	84	-0.01
29 S	Terphenyl-d14	0.972	0.870	10.5	86	-0.01
30	Benzo(a)anthracene	1.139	1.047	8.1	89	-0.01
31	Chrysene	1.193	1.157	3.0	93	-0.01
32	Bis(2-ethylhexyl)phthalate	2.313	2.101	9.2	87	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	0.774	34.8#	65	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	94	-0.01
35	Benzo(b)fluoranthene	1.094	1.097	-0.3	97	-0.01
36	Benzo(k)fluoranthene	1.274	1.452	-14.0	111	-0.02
37 C	Benzo(a)pyrene	1.129	1.131	-0.2	96	-0.02
38	Dibenzo(a,h)anthracene	1.063	0.830	21.9	75	-0.02
39	Benzo(a,h,i)perylene	1.071	0.657	38.7#	59	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121918\
Data File : BE098530.D
Acq On : 19 Dec 2018 17:15
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 19 23:44:05 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			