

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121919\
 Data File : BE100958.D
 Acq On : 19 Dec 2019 22:42
 Operator : JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE121919

Quant Time: Dec 20 01:21:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 20 01:12:26 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	1.524	1.547	-1.5	99	0.00
3	n-Nitrosodimethylamine	0.438	0.246	43.8#	57	-0.01
4 S	2-Fluorophenol	0.714	0.625	12.5	90	0.00
5 S	Phenol-d6	0.871	0.782	10.2	89	0.00
6	bis(2-Chloroethyl)ether	1.023	0.655	36.0#	97	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
8 S	Nitrobenzene-d5	0.211	0.187	11.4	108	-0.01
9	Naphthalene	4.320	4.559	-5.5	105	0.00
10	Hexachlorobutadiene	0.185	0.196	-5.9	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.602	0.652	-8.3	108	0.00
12	2-Methylnaphthalene	0.608	0.622	-2.3	106	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.098	0.093	5.1	119	0.00
15 S	2-Fluorobiphenyl	1.415	1.441	-1.8	107	0.00
16	Acenaphthylene	1.648	1.639	0.5	105	0.00
17	Acenaphthene	1.121	1.176	-4.9	108	0.00
18	Fluorene	1.418	1.441	-1.6	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4-Bromophenyl-phenylether	0.178	0.170	4.5	110	0.00
21	Hexachlorobenzene	0.203	0.210	-3.4	109	0.00
22	Pentachlorophenol	0.030	0.028	6.7	155#	0.01
23	Phenanthrene	1.043	1.052	-0.9	109	0.00
24	Anthracene	0.701	0.683	2.6	109	0.00
25 SURR	Fluoranthene-d10	5.007	5.012	-0.1	106	0.00
26	Fluoranthene	1.447	1.460	-0.9	108	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
28	Pvrene	1.430	1.565	-9.4	108	0.00
29 S	Terphenyl-d14	0.912	0.900	1.3	106	0.00
30	Benzo(a)anthracene	1.049	1.017	3.1	107	0.00
31	Chrysene	1.321	1.354	-2.5	105	0.00
32	Bis(2-ethylhexyl)phthalate	3.299	2.647	19.8	99	0.00
33	Indeno(1,2,3-cd)pyrene	1.383	1.538	-11.2	104	0.00
34 I	Perylene-d12	1.000	1.000	0.0	105	0.00
35	Benzo(b)fluoranthene	0.887	0.842	5.1	101	0.00
36	Benzo(k)fluoranthene	1.184	1.243	-5.0	108	0.00
37 C	Benzo(a)pyrene	1.010	1.055	-4.5	99	0.00
38	Dibenzo(a,h)anthracene	0.851	0.844	0.8	102	0.00
39	Benzo(a,h,i)perylene	1.052	1.125	-6.9	105	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121919\
Data File : BE100958.D
Acq On : 19 Dec 2019 22:42
Operator : JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE121919

Quant Time: Dec 20 01:21:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 20 01:12:26 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			