

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121919\
 Data File : BE100952.D
 Acq On : 19 Dec 2019 19:01
 Operator : JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Manual Integrations
 APPROVED

Sohil
 12/20/2019 4:26:54 PM

Quant Time: Dec 20 00:47:44 2019

Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121919.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Fri Dec 20 00:29:37 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	2819	0.40	ng	0.00
7) Naphthalene-d8	10.52	136	14019	0.40	ng	0.00
13) Acenaphthene-d10	14.39	164	8646	0.40	ng	0.00
19) Phenanthrene-d10	17.11	188	19540	0.40	ng	-0.01
27) Chrysene-d12	21.29	240	20444	0.40	ng	0.00
34) Perylene-d12	23.73	264	24007	0.40	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.35	112	8686	1.30	ng	0.00
5) Phenol-d6	6.93	99	12227	1.21	ng	-0.04
8) Nitrobenzene-d5	8.89	82	15268	2.29	ng	-0.03
11) 2-Methylnaphthalene-d10	12.12	152	34341	1.69	ng	0.00
14) 2,4,6-Tribromophenol	15.86	330	4444	4.62	ng	-0.01
15) 2-Fluorobiphenyl	13.00	172	51492	1.50	ng	-0.01
25) Fluoranthene-d10	19.14	212	397096	1.72	ng	0.00
29) Terphenyl-d14	19.75	244	79228	1.62	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.31	88	9818	1.910	ng	# 78
3) n-Nitrosodimethylamine	3.62	42	5529	1.322	ng	# 73
6) bis(2-Chloroethyl)ether	7.17	93	11625m	1.191	ng	
9) Naphthalene	10.57	128	244057	1.638	ng	99
10) Hexachlorobutadiene	10.86	225	9803	1.688	ng	99
12) 2-Methylnaphthalene	12.20	142	36715	1.578	ng	97
16) Acenaphthylene	14.10	152	60983	1.705	ng	99
17) Acenaphthene	14.44	154	40600	1.647	ng	99
18) Fluorene	15.42	166	52921	1.665	ng	99
20) 4-Bromophenyl-phenylether	16.32	248	15006	1.565	ng	# 88
21) Hexachlorobenzene	16.42	284	16037	1.540	ng	98
22) Pentachlorophenol	16.79	266	3051m	4.235	ng	
23) Phenanthrene	17.15	178	87107	1.472	ng	100
24) Anthracene	17.24	178	67068	1.355	ng	99
26) Fluoranthene	19.17	202	115972	1.642	ng	99
28) Pyrene	19.53	202	120299	1.811	ng	99
30) Benzo(a)anthracene	21.28	228	96187	1.596	ng	98
31) Chrysene	21.33	228	107046	1.512	ng	98
32) Bis(2-ethylhexyl)phthalate	21.22	149	267807	4.904	ng	100
33) Indeno(1,2,3-cd)pyrene	26.29	276	118624	1.923	ng	# 61
35) Benzo(b)fluoranthene	22.98	252	94687	1.441	ng	97
36) Benzo(k)fluoranthene	23.03	252	121757m	1.652	ng	
37) Benzo(a)pyrene	23.62	252	102815	1.807	ng	# 94
38) Dibenzo(a,h)anthracene	26.32	278	90760	1.603	ng	# 93
39) Benzo(g,h,i)perylene	27.08	276	106037	1.752	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121919\
Data File : BE100952.D
Acq On : 19 Dec 2019 19:01
Operator : JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:54 PM

Quant Time: Dec 20 00:47:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 20 00:29:37 2019
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.75 min Scan# 424
Delta R.T. -0.00 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Instrument :
BNA_E
ClientSampled :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
152	100		
150	155.6	116.7	175.1
115	71.7	55.7	83.5

Manual Integrations
APPROVED




Sohil
12/20/2019 4:26:54 PM




#2

1,4-Dioxane
Concen: 1.910 ng
RT: 3.31 min Scan# 39
Delta R.T. -0.00 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Tgt Ion	Ratio	Lower	Upper
88	100		
58	68.9	38.6	57.8#
43	29.0	20.5	30.7





#3

n-Nitrosodimethylamine

Concen: 1.322 ng

RT: 3.62 min Scan# 66

Delta R.T. -0.04 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion:	42	Resp:	5529
Ion	Ratio	Lower	Upper
42	100		
74	124.0	77.8	116.6#
44	4.1	0.0	0.0#

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:54 PM



#4

2-Fluorophenol

Concen: 1.298 ng




RT: 5.35 min Scan# 216




Delta R.T. -0.00 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Tgt Ion:	112	Resp:	8686
Ion	Ratio	Lower	Upper
112	100		
64	67.9	58.6	88.0
63	146.2	121.0	181.6





#5

Phenol-d6

Concen: 1.211 ng

RT: 6.93 min Scan# 353

Delta R.T. -0.04 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Tot Ion:	99	Resp:	12227
Ion	Ratio	Lower	Upper
99	100		
42	22.9	10.2	15.2#
71	37.1	0.0	0.0#

Instrument :





BNA_E

ClientSampleId :

SSTDICC1.6

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:54 PM



#6

bis(2-Chloroethyl)ether

Concen: 1.191 ng m


RT: 7.17 min Scan# 374


Delta R.T. -0.01 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Tgt Ion:	93	Resp:	11625
Ion	Ratio	Lower	Upper
93	100		
63	340.5	209.4	314.2#
95	37.5	28.6	42.8





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.52 min Scan# 643

Delta R.T. 0.00 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.6	14.4
54	34.5	36.2	54.4#
68	7.9	8.4	12.6#

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:54 PM



#8

Nitrobenzene-d5

Concen: 2.287 ng




RT: 8.89 min Scan# 517


Delta R.T. -0.03 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Tgt Ion	Ratio	Lower	Upper
82	100		
128	116.1	135.0	202.6#
54	205.5	182.6	274.0





#9

Naphthalene

Concen: 1.638 ng

RT: 10.57 min Scan# 647

Delta R.T. 0.00 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion:	128	Resp:	244057
Ion Ratio	Lower	Upper	
128	100		
129	2.8	2.5	3.7
127	3.2	2.7	4.1

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:54 PM



#10

Hexachlorobutadiene

Concen: 1.688 ng




RT: 10.86 min Scan# 669


Delta R.T. 0.00 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Tgt Ion:	225	Resp:	9803
Ion Ratio	Lower	Upper	
225	100		
223	63.1	49.4	74.2
227	63.5	50.3	75.5









#11
 2-Methylnaphthalene-d10
 Concen: 1.690 ng
 RT: 12.12 min Scan# 757
 Delta R.T. 0.00 min
 Lab File: BE100952.D
 Acq: 19 Dec 2019 19:01

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.0	22.4




Manual Integrations
 APPROVED


Sohil
 12/20/2019 4:26:54 PM



#12
 2-Methylnaphthalene
 Concen: 1.578 ng
 RT: 12.20 min Scan# 762
 Delta R.T. 0.00 min
 Lab File: BE100952.D
 Acq: 19 Dec 2019 19:01

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.1	67.8	101.8
115	38.0	31.7	47.5





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.39 min Scan# 914

Delta R.T. 0.00 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Instrument :

BNA_E


ClientSampleId :

SSTDICC1.6




Tot Ion:	164	Resp:	8646
Ion	Ratio	Lower	Upper
164	100		
162	102.7	81.6	122.4
160	47.1	38.5	57.7

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:54 PM



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E



#14

2,4,6-Tribromophenol

Concen: 4.621 ng


RT: 15.86 min Scan# 1018

Delta R.T. -0.01 min



Lab File: BE100952.D


Acq: 19 Dec 2019 19:01

Tgt Ion:	330	Resp:	4444
Ion	Ratio	Lower	Upper
330	100		
332	95.7	84.2	126.2
141	46.6	39.3	58.9



Abundance Ion 330.00 (329.70 to 330.70): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#15
2-Fluorobiphenyl
Concen: 1.501 ng
RT: 13.00 min Scan# 818
Delta R.T. -0.01 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
172	100		
171	36.1	27.5	41.3
170	24.8	19.8	29.6




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:54 PM



#16
Acenaphthylene
Concen: 1.705 ng
RT: 14.10 min Scan# 894
Delta R.T. 0.00 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.0	22.4
153	13.0	10.6	16.0









#17
Acenaphthene
Concen: 1.647 ng
RT: 14.44 min Scan# 918
Delta R.T. 0.00 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
154	100		
153	114.0	92.3	138.5
152	56.7	46.4	69.6




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:54 PM



#18
Fluorene
Concen: 1.665 ng
RT: 15.42 min Scan# 985
Delta R.T. 0.00 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Tgt Ion	Ratio	Lower	Upper
166	100		
165	97.2	79.0	118.4
167	13.5	10.6	16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.11 min Scan# 1111

Delta R.T. -0.01 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion:	188	Resp:	19540
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	15.5	9.5	14.3#

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:54 PM



#20

4-Bromophenyl-phenylether

Concen: 1.565 ng




RT: 16.32 min Scan# 1052


Delta R.T. 0.00 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Tgt Ion:	248	Resp:	15006
Ion Ratio	Lower	Upper	
248	100		
250	100.3	79.5	119.3
141	60.9	66.6	100.0#





#21

Hexachlorobenzene

Concen: 1.540 ng

RT: 16.42 min Scan# 1060

Delta R.T. 0.00 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Instrument :

BNA_E


ClientSampleId :

SSTDICC1.6




Tot Ion:	284	Resp:	16037
Ion Ratio		Lower	Upper
284	100		
142	42.9	33.8	50.6
249	34.9	26.7	40.1

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:54 PM



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E



#22

Pentachlorophenol

Concen: 4.235 ng m


RT: 16.79 min Scan# 1087

Delta R.T. -0.03 min



Lab File: BE100952.D


Acq: 19 Dec 2019 19:01

Tgt Ion:	266	Resp:	3051
Ion Ratio		Lower	Upper
266	100		
264	68.9	64.5	96.7
268	64.9	74.6	111.8#



Abundance Ion 266.00 (265.70 to 266.70): E
Ion 264.00 (263.70 to 264.70): E
Ion 268.00 (267.70 to 268.70): E





#23

Phenanthrene

Concen: 1.472 ng

RT: 17.15 min Scan# 1114

Delta R.T. -0.01 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion:178	Resp:	87107
Ion Ratio	Lower	Upper
178	100	
176	19.2	15.4 23.0
179	15.3	12.5 18.7

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:54 PM



#24

Anthracene

Concen: 1.355 ng




RT: 17.24 min Scan# 1121


Delta R.T. -0.01 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Tgt Ion:178	Resp:	67068
Ion Ratio	Lower	Upper
178	100	
176	18.9	15.7 23.5
179	15.2	12.0 18.0





#25

Fluoranthene-d10

Concen: 1.719 ng

RT: 19.14 min Scan# 1371

Delta R.T. 0.00 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Instrument :

BNA_E


ClientSampleId :

SSTDICC1.6

Tot Ion	Ratio	Resp	Lower	Upper
212	100	397096		
106	3.4		2.6	3.8
104	1.9		1.4	2.2

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:54 PM






Abundance

Ion 212.00 (211.70 to 212.70): E

Ion 106.00 (105.70 to 106.70): E

Ion 104.00 (103.70 to 104.70): E



#26

Fluoranthene

Concen: 1.642 ng


RT: 19.17 min Scan# 1376

Delta R.T. 0.00 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Tgt Ion	Ratio	Resp	Lower	Upper
202	100	115972		
101	12.1		9.0	13.6
203	16.7		13.4	20.0






Abundance

Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#27

Chrysene-d12

Concen: 0.400 ng

RT: 21.29 min Scan# 1624

Delta R.T. -0.00 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion:	240	Resp:	20444
Ion Ratio	Lower	Upper	
240	100		
120	12.2	11.5	17.3
236	27.9	22.8	34.2

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:54 PM



#28

Pyrene

Concen: 1.811 ng




RT: 19.53 min Scan# 1439


Delta R.T. 0.00 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Tgt Ion:	202	Resp:	120299
Ion Ratio	Lower	Upper	
202	100		
200	21.2	16.4	24.6
203	17.5	13.8	20.8









#29
 Terphenyl-d14
 Concen: 1.618 ng
 RT: 19.75 min Scan# 1476
 Delta R.T. -0.01 min
 Lab File: BE100952.D
 Acq: 19 Dec 2019 19:01

Instrument :
 BNA_E
 ClientSampled :
 SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
244	100		
212	32.4	30.2	45.2
122	12.9	10.7	16.1




Manual Integrations
 APPROVED


Sohil
 12/20/2019 4:26:54 PM



#30
 Benzo(a)anthracene
 Concen: 1.596 ng
 RT: 21.28 min Scan# 1623
 Delta R.T. -0.00 min
 Lab File: BE100952.D
 Acq: 19 Dec 2019 19:01

Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.0	21.9	32.9
229	20.1	15.9	23.9









#31
Chrysene
Concen: 1.512 ng
RT: 21.33 min Scan# 1627
Delta R.T. -0.00 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
228	100		
226	29.5	24.2	36.4
229	19.6	16.6	24.8




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:54 PM



#32
Bis(2-ethylhexyl)phthalate
Concen: 4.904 ng
RT: 21.22 min Scan# 1618
Delta R.T. -0.00 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Tgt Ion	Ratio	Lower	Upper
149	100		
167	6.4	5.0	7.6
279	0.9	0.7	1.1









#33
Indeno(1,2,3-cd)pyrene
Concen: 1.923 ng
RT: 26.29 min Scan# 2874
Delta R.T. -0.01 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
276	100		
138	23.4	9.0	13.4#
277	24.8	5.9	8.9#




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:54 PM



#34
Perylene-d12
Concen: 0.400 ng
RT: 23.73 min Scan# 2160
Delta R.T. -0.00 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.4	29.2
265	32.5	26.4	39.6









#35
Benzo(b)fluoranthene
Concen: 1.441 ng
RT: 22.98 min Scan# 1951
Delta R.T. -0.00 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.1	18.6	27.8
125	10.7	10.2	15.2




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:54 PM



#36
Benzo(k)fluoranthene
Concen: 1.652 ng m
RT: 23.03 min Scan# 1965
Delta R.T. -0.00 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.1	19.2	28.8
125	12.0	17.9	26.9#









#37
Benzo(a)pyrene
Concen: 1.807 ng
RT: 23.62 min Scan# 2130
Delta R.T. 0.00 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.2	19.9	29.9
125	12.2	12.3	18.5




Manual Integrations
APPROVED




Sohil
12/20/2019 4:26:54 PM



#38
Dibenzo(a,h)anthracene
Concen: 1.603 ng
RT: 26.32 min Scan# 2883
Delta R.T. -0.01 min
Lab File: BE100952.D
Acq: 19 Dec 2019 19:01

Tgt Ion	Ratio	Lower	Upper
278	100		
139	16.6	17.0	25.6
279	23.9	21.1	31.7





#39

Benzo(a,h,i)perylene

Concen: 1.752 ng

RT: 27.08 min Scan# 3095

Delta R.T. 0.00 min

Lab File: BE100952.D

Acq: 19 Dec 2019 19:01

Tot Ion:	276	Resp:	106037
Ion Ratio	Lower	Upper	
276	100		
277	23.5	19.9	29.9
138	22.0	18.4	27.6

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:54 PM

