

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121919\
 Data File : BE100954.D
 Acq On : 19 Dec 2019 20:13
 Operator : JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Manual Integrations
 APPROVED

Sohil
 12/20/2019 4:26:57 PM

Quant Time: Dec 20 01:15:13 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 20 00:29:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	2619	0.40	ng	0.00
7) Naphthalene-d8	10.52	136	13024	0.40	ng	0.00
13) Acenaphthene-d10	14.37	164	8277	0.40	ng	-0.02
19) Phenanthrene-d10	17.11	188	18689	0.40	ng	-0.01
27) Chrysene-d12	21.29	240	21721	0.40	ng	0.00
34) Perylene-d12	23.73	264	25032	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.35	112	27763	4.47	ng	0.00
5) Phenol-d6	0.00	99	0d	0.00	ng	
8) Nitrobenzene-d5	0.00	82	0d	0.00	ng	
11) 2-Methylnaphthalene-d10	12.12	152	100576	5.33	ng	0.00
14) 2,4,6-Tribromophenol	0.00	330	0d	0.00	ng	
15) 2-Fluorobiphenyl	13.00	172	147476	4.49	ng	-0.02
25) Fluoranthene-d10	19.14	212	1243142	5.63	ng	0.00
29) Terphenyl-d14	19.75	244	256614	4.93	ng	0.00
Target Compounds						
3) n-Nitrosodimethylamine	3.61	42	19452	5.005	ng	Qvalue # 76
6) bis(2-Chloroethyl)ether	7.17	93	40762m	4.496	ng	
9) Naphthalene	10.56	128	706866	5.108	ng	99
10) Hexachlorobutadiene	10.86	225	27186	5.038	ng	98
12) 2-Methylnaphthalene	12.19	142	110794	5.127	ng	97
16) Acenaphthylene	14.10	152	194477	5.680	ng	98
17) Acenaphthene	14.44	154	118837	5.036	ng	98
18) Fluorene	15.41	166	159516	5.242	ng	100
20) 4-Bromophenyl-phenylether	16.31	248	46377	5.058	ng	90
21) Hexachlorobenzene	16.43	284	47709	4.789	ng	98
23) Phenanthrene	17.15	178	266258	4.703	ng	100
26) Fluoranthene	19.17	202	364996	5.402	ng	99
28) Pyrene	19.53	202	377972	5.356	ng	99
30) Benzo(a)anthracene	21.28	228	349587	5.460	ng	99
31) Chrysene	21.33	228	352942	4.692	ng	99
32) Bis(2-ethylhexyl)phthalate	21.22	149	923584	15.917	ng	99
33) Indeno(1,2,3-cd)pyrene	26.28	276	407986	6.226	ng	# 58
35) Benzo(b)fluoranthene	22.98	252	328090	4.788	ng	97
36) Benzo(k)fluoranthene	23.03	252	404101m	5.258	ng	
37) Benzo(a)pyrene	23.62	252	338836	5.711	ng	# 93
38) Dibenzo(a,h)anthracene	26.31	278	325018	5.506	ng	# 92
39) Benzo(g,h,i)perylene	27.07	276	354914	5.623	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121919\
Data File : BE100954.D
Acq On : 19 Dec 2019 20:13
Operator : JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:57 PM

Quant Time: Dec 20 01:15:13 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 20 00:29:37 2019
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.75 min Scan# 424
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
152	100		
150	154.5	116.7	175.1
115	66.8	55.7	83.5

Manual Integrations
APPROVED




Sohil
12/20/2019 4:26:57 PM




#3

n-Nitrosodimethylamine
Concen: 5.005 ng
RT: 3.61 min Scan# 65
Delta R.T. -0.05 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Tgt Ion	Ratio	Lower	Upper
42	100		
74	121.0	77.8	116.6#
44	8.3	0.0	0.0#





#4

2-Fluorophenol

Concen: 4.467 ng

RT: 5.35 min Scan# 216

Delta R.T. -0.00 min

Lab File: BE100954.D

Acq: 19 Dec 2019 20:13

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tot Ion:	112	Resp:	27763
Ion Ratio	Lower	Upper	
112	100		
64	69.5	58.6	88.0
63	146.5	121.0	181.6

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:57 PM



#5

Phenol-d6


Concen: 0.000 ng


Expected RT: 6.97 min

Lab File: BE100954.D

Acq: 19 Dec 2019 20:13

Tgt Ion:	99
Sig	Exp Ratio
99	100
42	12.7
71	0.0





#6





bis(2-Chloroethyl)ether
Concen: 4.496 ng/ml
RT: 7.17 min Scan# 374
Delta R.T. -0.01 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
93	100		
63	332.5	209.4	314.2#
95	35.7	28.6	42.8

Manual Integrations
APPROVED




Sohil
12/20/2019 4:26:57 PM




#7

Naphthalene-d8
Concen: 0.400 ng
RT: 10.52 min Scan# 643
Delta R.T. 0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.9	9.6	14.4
54	30.9	36.2	54.4#
68	7.2	8.4	12.6#







#8
 Nitrobenzene-d5
 Concen: 0.000 ng
 Expected RT: 8.91 min
 Lab File: BE100954.D
 Acq: 19 Dec 2019 20:13

Tot Ion	82
Sig	Exp Ratio
82	100
128	168.8
54	228.3

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0




Manual Integrations
 APPROVED


Sohil
 12/20/2019 4:26:57 PM



#9
 Naphthalene
 Concen: 5.108 ng
 RT: 10.56 min Scan# 646
 Delta R.T. -0.01 min
 Lab File: BE100954.D
 Acq: 19 Dec 2019 20:13

Tgt Ion	128	Resp	706866
Ion Ratio	Lower	Upper	
128	100		
129	2.7	2.5	3.7
127	3.4	2.7	4.1






#10
Hexachlorobutadiene
Concen: 5.038 ng
RT: 10.86 min Scan# 669
Delta R.T. 0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
225	100		
223	62.9	49.4	74.2
227	64.1	50.3	75.5

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:57 PM






Abundance

Ion 225.00 (224.70 to 225.70): E


Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E



#11
2-Methylnaphthalene-d10
Concen: 5.329 ng
RT: 12.12 min Scan# 757
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13



Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.0	22.4




Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E






#12
2-Methylnaphthalene
Concen: 5.127 ng
RT: 12.19 min Scan# 762
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
142	100		
141	86.9	67.8	101.8
115	37.5	31.7	47.5

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:57 PM






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.37 min Scan# 913
Delta R.T. -0.02 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13




Tgt Ion	Ratio	Lower	Upper
164	100		
162	106.3	81.6	122.4
160	52.5	38.5	57.7


Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E







#14
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 15.88 min

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion: 330
Sig Exp Ratio
330 100
332 105.2
141 49.1




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:57 PM



#15
2-Fluorobiphenyl
Concen: 4.492 ng
RT: 13.00 min Scan# 818
Delta R.T. -0.02 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Tgt Ion: 172 Resp: 147476
Ion Ratio Lower Upper
172 100
171 35.5 27.5 41.3
170 24.3 19.8 29.6









#16
Acenaphthylene
Concen: 5.680 ng
RT: 14.10 min Scan# 894
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.0	22.4
153	13.0	10.6	16.0




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:57 PM



#17
Acenaphthene
Concen: 5.036 ng
RT: 14.44 min Scan# 918
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Tgt Ion	Ratio	Lower	Upper
154	100		
153	113.9	92.3	138.5
152	55.7	46.4	69.6









#18
Fluorene
 Concen: 5.242 ng
 RT: 15.41 min Scan# 984
 Delta R.T. -0.01 min
 Lab File: BE100954.D
 Acq: 19 Dec 2019 20:13

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tot Ion	Ratio	Resp	Lower	Upper
166	100			
165	98.9	79.0	118.4	
167	13.5	10.6	16.0	




Manual Integrations
 APPROVED


Sohil
 12/20/2019 4:26:57 PM



#19
Phenanthrene-d10
 Concen: 0.400 ng
 RT: 17.11 min Scan# 1111
 Delta R.T. -0.01 min
 Lab File: BE100954.D
 Acq: 19 Dec 2019 20:13

Tgt Ion	Ratio	Resp	Lower	Upper
188	100			
94	0.0	0.0	0.0	
80	13.6	9.5	14.3	









#20
4-Bromophenyl-phenylether
Concen: 5.058 ng
RT: 16.31 min Scan# 1051
Delta R.T. -0.01 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
248	100		
250	90.9	79.5	119.3
141	93.8	66.6	100.0




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:57 PM



#21
Hexachlorobenzene
Concen: 4.789 ng
RT: 16.43 min Scan# 1060
Delta R.T. 0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Tgt Ion	Ratio	Lower	Upper
284	100		
142	42.6	33.8	50.6
249	35.0	26.7	40.1





#23

Phenanthrene

Concen: 4.703 ng

RT: 17.15 min Scan# 1114

Delta R.T. -0.01 min

Lab File: BE100954.D

Acq: 19 Dec 2019 20:13

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tot Ion:178	Resp:	266258
Ion Ratio	Lower	Upper
178	100	
176	19.3	15.4 23.0
179	15.3	12.5 18.7

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:57 PM



#25

Fluoranthene-d10

Concen: 5.628 ng




RT: 19.14 min Scan# 1371


Delta R.T. -0.00 min

Lab File: BE100954.D

Acq: 19 Dec 2019 20:13

Tgt Ion:212	Resp:	1243142
Ion Ratio	Lower	Upper
212	100	
106	3.5	2.6 3.8
104	1.9	1.4 2.2









#26
Fluoranthene
Concen: 5.402 ng
RT: 19.17 min Scan# 1376
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Resp	Lower	Upper
202	100	364996		
101	12.2		9.0	13.6
203	17.0		13.4	20.0




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:57 PM



#27
Chrysene-d12
Concen: 0.400 ng
RT: 21.29 min Scan# 1624
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Tgt Ion	Ratio	Resp	Lower	Upper
240	100	21721		
120	9.2		11.5	17.3#
236	27.9		22.8	34.2









#28
Pvrene
Concen: 5.356 ng
RT: 19.53 min Scan# 1439
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion:	202	Resp:	377972
Ion Ratio	Lower	Upper	
202	100		
200	21.3	16.4	24.6
203	17.5	13.8	20.8




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:57 PM



#29
Terphenyl-d14
Concen: 4.932 ng
RT: 19.75 min Scan# 1476
Delta R.T. -0.01 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Tgt Ion:	244	Resp:	256614
Ion Ratio	Lower	Upper	
244	100		
212	31.6	30.2	45.2
122	12.0	10.7	16.1









#30
Benzo(a)anthracene
Concen: 5.460 ng
RT: 21.28 min Scan# 1623
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
228	100		
226	26.5	21.9	32.9
229	20.3	15.9	23.9




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:57 PM



#31
Chrysene
Concen: 4.692 ng
RT: 21.33 min Scan# 1627
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.0	24.2	36.4
229	19.9	16.6	24.8





#32

Bis(2-ethylhexyl)phthalate

Concen: 15.917 ng

RT: 21.22 min Scan# 1618

Delta R.T. -0.00 min

Lab File: BE100954.D

Acq: 19 Dec 2019 20:13

Instrument :

BNA_E


ClientSampleId :

SSTDICC5.0

Tot Ion:	149	Resp:	923584
Ion Ratio	Lower	Upper	
149	100		
167	6.5	5.0	7.6
279	0.8	0.7	1.1

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:57 PM






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#33

Indeno(1,2,3-cd)pyrene

Concen: 6.226 ng

RT: 26.28 min Scan# 2873

Delta R.T. -0.01 min

Lab File: BE100954.D

Acq: 19 Dec 2019 20:13




Tgt Ion:	276	Resp:	407986
Ion Ratio	Lower	Upper	
276	100		
138	24.8	9.0	13.4#
277	25.0	5.9	8.9#


Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#34
Pervlene-d12
Concen: 0.400 ng
RT: 23.73 min Scan# 2161
Delta R.T. 0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
264	100		
260	23.7	19.4	29.2
265	30.5	26.4	39.6




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:57 PM



#35
Benzo(b)fluoranthene
Concen: 4.788 ng
RT: 22.98 min Scan# 1951
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.1	18.6	27.8
125	10.6	10.2	15.2









#36
Benzo(k)fluoranthene
Concen: 5.258 ng
RT: 23.03 min Scan# 1965
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.2	19.2	28.8
125	11.1	17.9	26.9




Manual Integrations
APPROVED


Sohil
12/20/2019 4:26:57 PM



#37
Benzo(a)pyrene
Concen: 5.711 ng
RT: 23.62 min Scan# 2130
Delta R.T. 0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.2	19.9	29.9
125	11.4	12.3	18.5









#38
Dibenzo(a,h)anthracene
Concen: 5.506 ng
RT: 26.31 min Scan# 2881
Delta R.T. -0.02 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
278	100		
139	16.4	17.0	25.6#
279	23.6	21.1	31.7

Manual Integrations
APPROVED

Sohil
12/20/2019 4:26:57 PM



#39
Benzo(g,h,i)perylene
Concen: 5.623 ng
RT: 27.07 min Scan# 3093
Delta R.T. -0.00 min
Lab File: BE100954.D
Acq: 19 Dec 2019 20:13

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.5	19.9	29.9
138	22.2	18.4	27.6

