

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122018\
 Data File : BE098534.D
 Acq On : 20 Dec 2018 12:01
 Operator : SJ/JU
 Sample : J6319-18
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 063_SW-E-1

Quant Time: Dec 20 13:33:00 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	408	0.40	ng	-0.01
7) Naphthalene-d8	10.65	136	2012	0.40	ng	-0.01
13) Acenaphthene-d10	14.51	164	1904	0.40	ng	-0.02
19) Phenanthrene-d10	17.26	188	5631	0.40	ng	-0.02
27) Chrysene-d12	21.45	240	9380	0.40	ng	-0.01
34) Perylene-d12	23.98	264	8667	0.40	ng	-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.44	112	312	0.33	ng	0.00
5) Phenol-d6	7.04	99	263	0.19	ng	0.00
8) Nitrobenzene-d5	9.02	82	598	0.33	ng	0.00
11) 2-Methylnaphthalene-d10	12.25	152	1479	0.45	ng	-0.02
14) 2,4,6-Tribromophenol	16.01	330	461	0.66	ng	-0.01
15) 2-Fluorobiphenyl	13.13	172	2482	0.31	ng	-0.02
25) Fluoranthene-d10	19.29	212	38485	0.53	ng	-0.02
29) Terphenyl-d14	19.89	244	9278	0.41	ng	-0.02
Target Compounds						
9) Naphthalene	10.69	128	3609	0.178	ng	# 88
12) 2-Methylnaphthalene	12.32	142	330	0.100	ng	# 86
17) Acenaphthene	14.57	154	195	0.036	ng	99
23) Phenanthrene	17.30	178	553	0.040	ng	# 93
32) Bis(2-ethylhexyl)phthalate	21.35	149	7509	0.138	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122018\
Data File : BE098534.D
Acq On : 20 Dec 2018 12:01
Operator : SJ/JU
Sample : J6319-18
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
063_SW-E-1

Quant Time: Dec 20 13:33:00 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

