

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122018\
 Data File : BE098538.D
 Acq On : 20 Dec 2018 14:30
 Operator : SJ/JU
 Sample : J6340-12
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 TWP-01

Quant Time: Dec 21 05:02:37 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	34197	0.40	ng	-0.02
7) Naphthalene-d8	10.64	136	150945	0.40	ng	-0.03
13) Acenaphthene-d10	14.52	164	115725	0.40	ng	-0.01
19) Phenanthrene-d10	17.26	188	369739	0.40	ng	-0.02
27) Chrysene-d12	21.45	240	561915	0.40	ng	-0.01
34) Perylene-d12	23.99	264	516106	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.43	112	144955	1.81	ng	-0.01
5) Phenol-d6	7.01	99	143199	1.26	ng	-0.03
8) Nitrobenzene-d5	9.00	82	360488	2.62	ng	-0.01
11) 2-Methylnaphthalene-d10	12.25	152	10	0.00	ng	-0.01
14) 2,4,6-Tribromophenol	16.01	330	245005	5.77	ng	-0.01
15) 2-Fluorobiphenyl	13.13	172	795336	1.63	ng	-0.01
25) Fluoranthene-d10	19.30	212	137	0.00	ng	-0.01
29) Terphenyl-d14	19.89	244	2372931	1.74	ng	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122018\
Data File : BE098538.D
Acq On : 20 Dec 2018 14:30
Operator : SJ/JU
Sample : J6340-12
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
TWP-01

Quant Time: Dec 21 05:02:37 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

